

INTERNATIONAL STANDARD

IEC 61675-2

Edition 1.1

2005-02

Edition 1:1998 consolidated with amendment 1:2004

Radionuclide imaging devices – Characteristics and test conditions –

Part 2: Single photon emission computed tomographs

© IEC 2005 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

CJ

For price, see current catalogue

CONTENTS

FOREWORD	3
1 General.....	5
1.1 Scope and object.....	5
1.2 Normative references	5
2 Terminology and definitions	6
3 Test methods	12
3.1 Calibration measurements	13
3.2 Measurement of COLLIMATOR hole misalignment	14
3.3 Measurement of SPECT system SENSITIVITY	14
3.4 Scatter	16
3.5 Measurement of SPECT non-uniformity of response.....	18
3.6 SPECT system SPATIAL RESOLUTION.....	18
3.7 Test methods for single photon computer tomographs operated in coincidence detection mode	19
4 ACCOMPANYING DOCUMENTS.....	32
 Annex A Index of defined terms	 46
 Figure 1 – Geometry of PROJECTIONS.....	 35
Figure 2 – Cylindrical head phantom	36
Figure 3 – Phantom insert with holders for the scatter source.....	37
Figure 4 – Evaluation of SCATTER FRACTION	38
Figure 5 – Reporting TRANSVERSE RESOLUTION	39
Figure 6 – Evaluation of FWHM	40
Figure 7 – Evaluation of EQUIVALENT WIDTH (EW)	41
Figure 8 – Phantom insert with hollow spheres.....	42
Figure 9 – Cross-section of body phantom	43
Figure 10 – Arm phantom	43
Figure 11 – Phantom configuration for COUNT RATE measurements according to 3.7.5.3.1.2 ..	44
Figure 12 – Scheme of the evaluation of COUNT LOSS correction	44
Figure 13 – Phantom insert for the evaluation of ATTENUATION correction	45

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**RADIONUCLIDE IMAGING DEVICES –
CHARACTERISTICS AND TEST CONDITIONS –****Part 2: Single photon emission computed tomographs**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61675-2 has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

This consolidated version of IEC 61675-2 consists of the first edition (1998) [documents 62C/206/FDIS and 62C/215/RVD] and its amendment 1 (2004) [documents 62C/378/FDIS and 62C/379/RVD].

The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience.

It bears the edition number 1.1.

A vertical line in the margin shows where the base publication has been modified by amendment 1.

In this standard, the following print types are used:

- TERMS DEFINED IN CLAUSE 2 OF THIS STANDARD OR LISTED IN ANNEX A: SMALL CAPITALS.

The requirements are followed by specifications for the relevant tests.

Annex A is for information only.

A bilingual version of this standard may be issued at a later date.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

RADIONUCLIDE IMAGING DEVICES – CHARACTERISTICS AND TEST CONDITIONS –

Part 2: Single photon emission computed tomographs

1 General

1.1 Scope and object

This part of IEC 61675 specifies terminology and test methods for describing the characteristics of Anger type rotational GAMMA CAMERA SINGLE PHOTON EMISSION COMPUTED TOMOGRAPHS (SPECT), equipped with parallel hole collimators. As these systems are based on Anger type GAMMA CAMERAS this part of IEC 61675 shall be used in conjunction with IEC 60789. These systems consist of a gantry system, single or multiple DETECTOR HEADS and a computer system together with acquisition, recording, and display devices.

This part of IEC 61675-2 also specifies test conditions for declaring the characteristics of single photon computer tomographs operated in coincidence mode as well as in single photon mode.

The test methods specified for coincidence mode are based on the test methods for dedicated PET tomographs as described in IEC 61675-1 to reflect as well as possible the clinical use of coincidence detection. Tests have been modified to reflect the limited sensitivity and COUNT RATE CHARACTERISTICS of the single photon computer tomographs operated in coincidence detection mode only when needed.

The test methods specified in this part of IEC 61675 have been selected to reflect as much as possible the clinical use of Anger type rotational GAMMA CAMERA SINGLE PHOTON EMISSION COMPUTED TOMOGRAPHS (SPECT). It is intended that the test methods be carried out by manufacturers thereby enabling them to describe the characteristics of SPECT systems on a common basis.

No test has been specified to characterize the uniformity of reconstructed images because all methods known so far will mostly reflect the noise of the image.

1.2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60788:1984, *Medical radiology – Terminology*

IEC 60789:1992, *Characteristics and test conditions of radionuclide imaging devices – Anger type gamma cameras*

IEC 61675-1, — *Radionuclide imaging devices – Characteristics and test conditions – Part 1: Positron emission tomographs*