

Technical Information Report

AAMI TIR12:2004

Designing, testing, and labeling reusable medical devices for reprocessing in health care facilities: A guide for medical device manufacturers

Designing, testing, and labeling reusable medical devices for reprocessing in health care facilities: A guide for medical device manufacturers

Approved 23 December 2004
Association for the Advancement of Medical Instrumentation

Abstract: This technical information report (TIR) covers design considerations that medical device manufacturers should take into account to help ensure that their products can be safely and effectively reprocessed. It also provides information on decontamination, disinfection, and sterilization processes commonly used in health care facilities so that manufacturers can validate reprocessing procedures that can be recommended to and performed adequately in health care facilities. Labeling recommendations and information on applicable regulations are also provided in the TIR, as well as a bibliography and other informative annexes.

Keywords: cleaning, decontamination, disinfection, instructions for use, medical device design, sterilization

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field of technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

Published by

Association for the Advancement of Medical Instrumentation
1110 N. Glebe Road, Suite 220
Arlington, VA 22201-4795

© 2005 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-230-3

Contents

	Page
Glossary of equivalent standards	v
Committee representation	vii
Foreword	x
Introduction	1
1 Scope	1
1.1 Inclusions	1
1.2 Exclusions	2
2 Definitions and abbreviations	2
3 Design considerations	4
3.1 General	4
3.2 Categories of medical devices	5
3.3 Physical design considerations	6
3.4 Material design considerations	7
3.4.1 General considerations	7
3.4.2 Polymeric materials	7
3.4.3 Metals	7
3.5 Total system design considerations	8
3.6 Use- and misuse-related design considerations	8
3.7 Device equivalence	8
3.7.1 General considerations	8
3.7.2 Product families	8
3.7.3 Master products	8
3.7.4 Equivalence assessment	9
4 Decontamination	9
4.1 General	9
4.2 Common hospital cleaning/decontamination agents and procedures	10
4.2.1 General considerations	10
4.2.2 Preparation at the point of use	10
4.2.3 Cleaning agents	10
4.2.4 Cleaning methods and equipment	12
4.3 Manufacturers' responsibilities	13
4.3.1 General considerations	13
4.3.2 Water quality for cleaning	14
4.3.3 Cleaning agents	14
4.3.4 Cleaning procedures	14
4.3.5 Test data and user verification	15
5 Disinfection with liquid chemicals	15
5.1 General	15
5.2 Levels of disinfectant activity	16
5.2.1 General considerations	16
5.2.2 High-level disinfection	16
5.2.3 Intermediate-level disinfection	18
5.2.4 Low-level disinfection	18
5.3 Device design considerations for disinfection with liquid chemicals	19
5.4 Criteria for selecting an appropriate chemical disinfectant	19
5.5 Toxicity	20
5.6 Materials compatibility	20
5.7 Efficacy of the process	20
5.8 Test data and user verification	20

6	Sterilization	20
6.1	General	20
6.2	Sterilization processes available for use in health care facilities	21
6.3	Device design considerations for sterilization	21
6.4	Packaging considerations for sterilization processes	22
6.5	Sterilization efficacy testing	23
6.6	Device and sterilization compatibility	24
6.7	Evaluation of sterilant residues and aeration or rinsing parameters	24
6.7.1	General considerations	24
6.7.2	Ethylene oxide	24
6.7.3	Other sterilant residues	24
6.7.4	Rinsing parameters	25
6.8	Documentation of qualification	25
6.9	Information to be supplied to health care personnel	26

Annexes

A	Liquid chemical disinfectants commonly used in health care facilities	27
B	Sterilization cycles commonly available in health care facilities	29
C	Processing CJD-contaminated patient care equipment and environmental surfaces	33
D	Regulatory considerations	34
E	Bibliography	39

Tables

1	Microorganisms listed in descending order of resistance to chemical sterilants and disinfectants	17
2	Levels of disinfection according to type of microorganism	18
3	Labeled contact conditions for high-level disinfection for FDA-cleared glutaraldehyde products	18
B.1	Parameters for gravity-displacement steam sterilization cycles in health care facilities	29
B.2	Parameters for dynamic-air-removal steam sterilization cycles in health care facilities	30
B.3	Parameters for EO sterilization cycles in health care facilities	30
B.4	Parameters for dry heat sterilization cycles in health care facilities	31
B.5	Parameters for liquid chemical sterilant cycles in health care facilities	31
B.6	Parameters for hydrogen peroxide gas plasma sterilization cycles in health care facilities	32
B.7	Parameters for chemical vapor sterilization cycles in health care facilities	32
B.8	Parameters for ozone sterilization cycles in health care facilities	32

Glossary of equivalent standards

International standards adopted in the United States may include normative references to other international standards. For each international standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the international standard. (Note: Documents are sorted by international designation.)

Other normatively referenced international standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1-2:2001 and Amendment 1:2004	ANSI/AAMI/IEC 60601-1-2:2001 and Amendment 1:2004	Identical
IEC 60601-2-04:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:1990 and Amendment 1:1996	ANSI/AAMI I136:2004	Major technical variations
IEC 60601-2-20:1990 and Amendment 1:1996	ANSI/AAMI I151:2004	Major technical variations
IEC 60601-2-21:1994 and Amendment 1:1996	ANSI/AAMI/IEC 60601-2-21 and Amendment 1:2000 (consolidated texts)	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004	Major technical variations
IEC TR 60878:2003	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC TR 62296:2003	ANSI/AAMI/IEC TIR62296:2003	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:1996	ANSI/AAMI/ISO 7199:1996/(R)2002	Identical
ISO 10993-1:2003	ANSI/AAMI/ISO 10993-1:2003	Identical
ISO 10993-2:1992	ANSI/AAMI/ISO 10993-2:1993/(R)2001	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003	Identical
ISO 10993-4:2002	ANSI/AAMI/ISO 10993-4:2002	Identical
ISO 10993-5:1999	ANSI/AAMI/ISO 10993-5:1999	Identical
ISO 10993-6:1994	ANSI/AAMI/ISO 10993-6:1995/(R)2001	Identical
ISO 10993-7:1995	ANSI/AAMI/ISO 10993-7:1995/(R)2001	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999	Identical
ISO 10993-10:2002	ANSI/AAMI BE78:2002	Minor technical variations
ISO 10993-11:1993	ANSI/AAMI 10993-11:1993	Minor technical variations
ISO 10993-12:2002	ANSI/AAMI/ISO 10993-12:2002	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2003	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002	Identical

International designation	U.S. designation	Equivalency
ISO 11134:1994	ANSI/AAMI/ISO 11134:1993	Identical
ISO 11135:1994	ANSI/AAMI/ISO 11135:1994	Identical
ISO 11137:1995 and Amdt 1:2001	ANSI/AAMI/ISO 11137:1994 and A1:2002	Identical
ISO 11138-1:1994	ANSI/AAMI ST59:1999	Major technical variations
ISO 11138-2:1994	ANSI/AAMI ST21:1999	Major technical variations
ISO 11138-3:1995	ANSI/AAMI ST19:1999	Major technical variations
ISO TS 11139:2001	ANSI/AAMI/ISO 11139:2002	Identical
ISO 11140-1:1995 and Technical Corrigendum 1:1998	ANSI/AAMI ST60:1996	Major technical variations
ISO 11140-5:2000	ANSI/AAMI ST66:1999	Major technical variations
ISO 11607:2003	ANSI/AAMI/ISO 11607:2000	Identical
ISO 11737-1:1995	ANSI/AAMI/ISO 11737-1:1995	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO 11737-3:2004	ANSI/AAMI/ISO 11737-3:2004	Identical
ISO TR 13409:1996	AAMI/ISO TIR13409:1996	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003	Identical
ISO 13488:1996	ANSI/AAMI/ISO 13488:1996	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998	Identical
ISO 14161:2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2000 and A1:2003	ANSI/AAMI/ISO 14971:2000 and A1:2003	Identical
ISO 15223:2000, A1:2002, and A2:2004	ANSI/AAMI/ISO 15223:2000, A1:2001, and A2:2004	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000 and A1:2004	Identical
ISO 15674:2001	ANSI/AAMI/ISO 15674:2001	Identical
ISO 15675:2001	ANSI/AAMI/ISO 15675:2001	Identical
ISO TS 15843:2000	ANSI/AAMI/ISO TIR15843:2000	Identical
ISO TR 15844:1998	AAMI/ISO TIR15844:1998	Identical
ISO 15882:2003	ANSI/AAMI/ISO 15882:2003	Identical
ISO TR 16142:1999	ANSI/AAMI/ISO TIR16142:2000	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 25539-1:2003	ANSI/AAMI/ISO 25539-1:2003	Identical

Committee representation

Association for the Advancement of Medical Instrumentation AAMI Instructions for Reusable Device Reprocessing Working Group

This technical information report was developed by the AAMI Instructions for Reusable Device Reprocessing Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the TIR does not necessarily mean that all working group members voted for its approval.

At the time this document was published, the **AAMI Instructions for Reusable Device Reprocessing Working Group** had the following members:

Cochairs: Nancy Chobin, RN, CSPDM
Sandra A. Lee, RN

Members: Michelle J. Alfa, PhD, FCCM, St. Boniface General Hospital, Winnipeg, Manitoba, Canada
Shelley Allen, AppTec Laboratory Services
Chris B. Anderson, St. Jude Medical Inc.
Lisa Baryschpolec, Ethicon Inc./Johnson & Johnson
Carl W. Bruch, PhD, Hudson, WI
Virginia C. Chamberlain, PhD, Hendersonville, NC
Nancy Chobin, RN, CSPDM, Saint Barnabas Health Care System, West Orange, NJ
Anne M. Cofiell, CRCST, International Association of Healthcare Central Service
Materiel Management
Jacqueline Daley, Sinai Hospital of Baltimore
Anthony J. DeMarinis, MS, CQA, CQM, Davol Inc./CR Bard
Michele Dawn DeMeo, Holy Name Hospital, Teaneck, NJ
Sue Ellen Erickson, MS, CNOR, Newart Beth Israel Medical Center, Cranford, NJ
LeRoy J. Fischbach, Minntech Corporation
France Frechette, MSc, TSO₃ Inc.
Gregory G. Ginsberg, MD, American Society for Gastrointestinal Endoscopy
Zory R. Glaser, PhD, MPH, CSPDM, Johns Hopkins University School of Public Health
Steve N. Goldstine, PhD, DDS, Albuquerque, NM
Wayne D. Grinthal, Olympus America Inc.
Douglas F. Harbrecht, Boston Scientific Corporation
Ronald J. Haselhorst, Cardinal Health
Steven M. Kovach, Healthmark Industries
Colleen Patricia Landers, RN, Canadian Standards Association
Sandra A. Lee, RN, STERIS Corporation
Chiu Lin, PhD, U.S. Food and Drug Administration
Teckla A. Maresca, LPN, CSPDM, St. Clare's Health System, Denville, NJ
Patrick J. McCormick, PhD, Bausch & Lomb Inc.
Russell D. Mills, Zimmer Inc.
Emily Mitzel, MS, Nelson Laboratories Inc.
David Parente, NAMSA
Leslie Phillips, STS duoTEK
Janet Prust, 3M Health Care
Timothy Ramsey, Northview Biosciences
Marimargaret Reichert, RN, MA, Southwest General Health Center, Middleburg Heights, OH
Linda Slone, RN, CNOR, Sibley Memorial Hospital, Washington, DC
Eugenia Stern, MBA, MS, Alcon Laboratories Inc.
David B. Vogel, Smith & Nephew, Inc.
Nora Wikander, RN, CSPDM, St. Joseph's Wayne Hospital, Wayne, NJ
Christine H. Yunker, Abbott Laboratories

Alternates: Solomon Alade, PhD, Alcon Laboratories Inc.
Ralph J. Basile, MBA, Healthmark Industries
Richard H. Bean, Zimmer Inc.
John Broad, NAMSA
Trabue D. Bryans, AppTec Laboratory Services
Anne Burke, TSO₃ Inc.
Carlos Chavez, PhD, Abbott Laboratories
Charles Cogdill, Boston Scientific Corporation
Kimbrell Darnell, CR Bard
Thomas J. Frazar, DePuy/Johnson & Johnson

Jim Kaiser, Bausch & Lomb Inc.
Susan G. Klacik, ACE, International Association of Healthcare Central Service Materiel Management
Martha T. O'Lone, U.S. Food and Drug Administration
Sara Toole, RM, Nelson Laboratories Inc.
David Tucholski, Cardinal Health
John W. Walker, STERIS Corporation
Martha Young, 3M Health Care

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

AAMI Sterilization Standards Committee

Cochairs: Victoria M. Hitchins, PhD
William E. Young

Members: Trabue D. Bryans, AppTec Laboratory Services
Virginia C. Chamberlain, PhD, Hendersonville, NC
Nancy Chobin, RN, CSPDM, Saint Barnabas Health Care System, West Orange, NJ
Anne M. Cofield, CRCST, International Association of Healthcare Central Service Materiel Management
Charles Cogdill, Boston Scientific Corporation
Ramona Conner, RN, Association of periOperative Registered Nurses
Kimbrell Darnell, CR Bard
Georgia P. Dash, RN, MS, CIC, Association for Professionals in Infection Control and Epidemiology
Lisa Foster, Sterigenics International
James M. Gibson, Jr., JM Gibson Associates
Barbara J. Goodman, RN, CNOR, Rising Sun, MD
Joel R. Gorski, PhD, NAMSA
Deborah A. Havlik, Hospira Inc.
Victoria M. Hitchins, PhD, U.S. Food and Drug Administration
Clark W. Houghtling, Cosmed Group Inc.
Richard M. Johnson, MSc, Abbott Laboratories
Lois Atkinson Jones, MS, Cary, NC
Byron J. Lambert, PhD, Guidant Corporation
Colleen Patricia Landers, RN, Canadian Standards Association
Sandra A. Lee, RN, STERIS Corporation
David Liu, Johnson & Johnson
Jeff Martin, Alcon Laboratories Inc.
Patrick J. McCormick, PhD, Bausch & Lomb Inc.
Thomas K. Moore, Getinge USA
Barry F. J. Page, Garner, NC
Phil M. Schneider, 3M Health Care
Michael H. Scholla, MS, PhD, DuPont Nonwovens
Frank Sizemore, American Society for Healthcare Central Service Professionals
Gregory O. Stecklein, MS, MSM, Cardinal Health
William N. Thompson, TYCO Healthcare/Kendall
James Whitbourne, STS duoTEK
James L. Whitby, MA, MB, FRCP, London, Ontario, Canada
Thelma Wilcott, Becton Dickinson & Company
Martell Kress Winters, Nelson Laboratories Inc.
William E. Young, Deer Park, IL

Alternates: Lloyd Brown, TYCO Healthcare/Kendall
Lina C. Bueno, DuPont Nonwovens
Craig M. Herring, Johnson & Johnson
Jim Kaiser, Bausch & Lomb Inc.
Susan G. Klacik, ACE, International Association of Healthcare Central Service Materiel Management
Joseph J. Lasich, Alcon Laboratories Inc.
Chiu Lin, PhD, U.S. Food and Drug Administration
Lisa N. Macdonald, Becton Dickinson & Company
Ralph Makinen, Guidant Corporation
Mary S. Mayo, CR Bard

David Ford McGoldrick, Abbott Laboratories
Jerry R. Nelson, MS, PhD, Nelson Laboratories Inc.
Jeff Peltier, Boston Scientific Corporation
Janet Prust, 3M Health Care
Nancy J. Rakiewicz, STS duoTEK
Ralph Stick, AppTec Laboratory Services
John W. Walker, STERIS Corporation
Valerie Welter, Hospira Inc.
William T. Young, Sterigenics International

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

Foreword

This AAMI Technical Information Report (TIR) was developed by the AAMI Instructions for Device Reprocessing Working Group under the auspices of the AAMI Sterilization Standards Committee.

The first edition of this TIR was published in 1994. This edition, the second, provides up-to-date information on cleaning processes, cleaning verification, and currently available sterilization technologies. Also, the text and reference material have been generally updated for currency.

Suggestions for improving this TIR are invited. Comments and suggested revisions should be sent to Technical Programs, AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

Designing, testing, and labeling reusable medical devices for reprocessing in health care facilities: A guide for device manufacturers

Introduction

Scientific advances in diagnostic and therapeutic medicine have led to the development of new and sophisticated reusable medical devices and instruments for use by health care practitioners. These devices vary in size, complexity, fragility, immersibility, and sensitivity to cleaning, disinfecting, and sterilizing agents and processes. Manufacturers of reusable medical devices have the responsibility to support product label claims of reusability by providing complete and comprehensive written instructions for the handling, cleaning, disinfection, testing, packaging, sterilization, and, if applicable, aeration of their products. Manufacturers also have the responsibility to conduct and document any testing necessary to validate the suitability of these instructions. Manufacturers have these obligations under U.S. Food and Drug Administration (FDA) labeling regulations (21 CFR 801). Detailed FDA recommendations are provided in the FDA guidance document, *Labeling reusable medical devices for reprocessing in health care facilities: FDA reviewer guidance* (FDA, 1996; <http://www.fda.gov/cdrh/ode/198.pdf>).

Health care personnel have the responsibility to obtain and review manufacturers' data and recommendations and to ensure that they have the necessary resources to follow manufacturers' instructions thoroughly.

This TIR is intended to assist medical device manufacturers in the design, testing, and labeling of devices intended for reuse and reprocessing in health care facilities. Device manufacturers may wish to reassess the labeling of existing products in light of the recommendations of this TIR.

Secondarily, this TIR can also be a resource in identifying the questions health care professionals should ask manufacturers when considering a product for purchase or when devising a reprocessing protocol for a product already being used. See also AAMI TIR7.

NOTE—At this writing, AAMI TIR7 was under revision. A combined revision of AAMI TIR7 and ANSI/AAMI ST58 will ultimately be issued as ANSI/AAMI ST58, *Chemical sterilization and high-level disinfection in health care facilities*.

1 Scope

1.1 Inclusions

The scope of this TIR includes the following topics:

- a) *Design considerations*: Assurance that a device can be safely and effectively reprocessed begins with the design of the device. Section 3 of the TIR describes categories of medical devices and the materials and other design characteristics that affect the ability of health care personnel to clean, disinfect, and/or sterilize devices adequately.
- b) *Decontamination*: A device cannot be disinfected adequately or sterilized to an adequate sterility assurance level (SAL) if it cannot be cleaned thoroughly. Section 4 addresses variables associated with cleaning and other decontamination processes used in health care facilities, as well as the minimum information that the device manufacturer should supply to health care personnel.
- c) *Disinfection*: Section 5 describes the levels of disinfection, the criteria for selecting chemical disinfectants, and the testing that device manufacturers should perform to establish the effectiveness of the disinfection processes recommended for their products.
- d) *Sterilization*: Section 6 describes the sterilization processes commonly used in health care facilities, the minimum information that device manufacturers should provide with their products, and the procedures that device manufacturers should use to qualify the sterilization parameters that they recommend for their products.

This TIR also includes definitions of terms, a list of references, and annexes providing supplementary information.

1.2 Exclusions

This TIR does *not* cover the following topics:

- a) the design, testing, and labeling of reusable textiles (see ANSI/AAMI PB70) or
- b) the design, testing, and labeling of devices intended and labeled for single use.

Although this TIR refers to specifying water quality for cleaning and other elements of reprocessing, it does not address methods of ensuring adequate water quality. A TIR on defining, achieving, and maintaining appropriate water quality for reprocessing medical devices is under development by AAMI.

2 Definitions and abbreviations

2.1 aerator: Machine designed to speed the removal of ethylene oxide (EO) residuals from sterilized items by subjecting sterilized items to warm, circulating, filtered air.

2.2 antigen: According to *Dorland's Pocket Medical Dictionary* (1982), "any substance capable, under appropriate conditions, of inducing a specific immune response and of reacting with the products of that response . . . Antigens may be soluble substances, such as toxins and foreign proteins, or particulate, such as bacteria and tissue cells . . ."

2.3 bioburden: Population of viable microorganisms on a product and/or a package.

NOTE—When measured, bioburden is expressed as the total count of bacterial and fungal colony-forming units (CFUs) per single item.

2.4 biological indicator (BI): Microbiological test system providing a defined resistance to a specified sterilization process.

NOTE—BIs are intended to demonstrate whether the conditions were adequate to achieve sterilization. A negative BI does not prove that all items in the load are sterile or that all were exposed to adequate sterilization conditions.

2.5 CDC: Centers for Disease Control and Prevention.

2.6 chemical germicide: Disinfectant or sterilant that is labeled to disinfect or sterilize a medical device.

2.7 chemical indicator (CI): Device used to monitor the presence or attainment of one or more of the parameters required for a satisfactory sterilization process, or used in specific tests of sterilization equipment.

NOTE 1—ANSI/AAMI ST60:1996, *Sterilization of health care products—Chemical indicators—Part 1: General requirements*, defines five classes of CIs and specifies performance requirements for them:

Class 1 (process indicator): Chemical indicator intended for use with individual units (e.g., packs, containers) to demonstrate that the unit has been exposed to the sterilization process and to distinguish between processed and unprocessed units.

Class 2 (Bowie-Dick test indicator): Chemical indicator designed for use in a specific test procedure (e.g., the Bowie-Dick test used to determine if air removal has been adequate in a steam sterilization process).

Class 3 (single-parameter indicator): Chemical indicator designed to react to one of the critical parameters of sterilization and to indicate exposure to a sterilization cycle at a stated value of the chosen parameter.

Class 4 (multi-parameter indicator): Chemical indicator designed to react to two or more of the critical parameters of sterilization and to indicate exposure to a sterilization cycle at stated values of the chosen parameters.

Class 5 (integrating indicator): Chemical indicator designed to react to all critical parameters over a specified range of sterilization cycles and whose performance has been correlated to the performance of the stated test organism under the labeled conditions of use.

NOTE 2—CIs assist in the detection of potential sterilization failures that could result from incorrect packaging, incorrect loading of the sterilizer, or sterilizer malfunctions. The "pass" response of a CI does not prove that the item monitored by the indicator is sterile.

2.8 chemical vapor sterilization: Specific sterilization process that uses a solution of alcohol, water, and inert ingredients, with trace formaldehyde (less than 0.25 %); the solution is heated to produce an unsaturated vapor with temperature, pressure, and exposure time within specified limits.

NOTE—Although "chemical vapor" can be taken to refer to gaseous chemical sterilants in general, the term is used here, as it is commonly used in the health care community, to refer to a specific process.