

Technical Information Report

AAMI TIR15: 2009/(R)2013

Physical aspects of ethylene
oxide sterilization

Physical aspects of ethylene oxide sterilization

Approved 3 September 2009 and reaffirmed 20 May 2013 by
Association for the Advancement of Medical Instrumentation

Abstract: This technical information report (TIR) provides additional guidance to augment the ANSI/AAMI/ISO 11135 series. Topics covered in this TIR are sterilization equipment, considerations for preconditioning, calculations for relative humidity, calculations of ethylene oxide concentration, and flammability.

Keywords: sterilization equipment, ethylene oxide, validation, relative humidity, flammability, preconditioning

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

This document is not an American National Standard, and the material contained herein is not normative in nature.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about 5 years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

Published by

Association for the Advancement of Medical Instrumentation
1110 N. Glebe Road, Suite 220
Arlington, VA 22201-4795
www.aami.org

© 2010 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI at 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795. Phone: (703) 525-4890; Fax: (703) 276-0793.

Printed in the United States of America

ISBN 1-57020-367-9

Contents

	Page
Glossary of equivalent standards	iv
Committee representation	vi
Foreword	ix
1 Scope.....	1
2 Definitions	1
3 EO sterilization processing equipment.....	1
4 Determination of minimum product temperature prior to preconditioning.....	6
5 Calculation of moisture content.....	7
6 Calculation of EO concentration.....	10
7 Calculations for use of flammability curve	15
8 Guide for statistical comparison to determine process equivalence.....	20
Bibliography	24
Figure	
1 EO Cycle Flammability Diagram.....	16
Tables	
1 List of constants and molecular weights of some common EO and diluent combinations.....	10
2 Gas constants (R)	11
3 Properties of saturated steam—Commonly used temperatures.....	14
4 Flammability review, partial pressures, and volume percentage	18

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005 Technical Corrigendum 1 and 2	ANSI/AAMI ES60601-1:2005 ANSI/AAMI ES60601-1:2005/C1:2009 (amdt)	Major technical variations C1 Identical to Corrigendum 1 & 2
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2009	ANSI/AAMI/IEC 60601-2-2:2009	Identical
IEC 60601-2-4:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:2009	ANSI/AAMI/IEC 60601-2-19:2009	Identical
IEC 60601-2-20:2009	ANSI/AAMI/IEC 60601-2-20:2009	Identical
IEC 60601-2-21:2009	ANSI/AAMI/IEC 60601-2-21:2009	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004/(R)2009	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2009	ANSI/AAMI/IEC 60601-2-50:2009	Identical
IEC 80601-2-30:2009 and Technical Corrigendum 1	ANSI/AAMI/IEC 80601-2-30:2009 and ANSI/AAMI/IEC 80601-2-30:2009/ C1:2009 (amdt) – consolidated text	Identical (with inclusion) C1 Identical to Corrigendum 1
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2009	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2009	ANSI/AAMI/IEC TIR62296:2009	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
IEC/TR 62354:2009	ANSI/AAMI/IEC TIR62354:2009	Identical
IEC/TR 80002-1:2009	ANSI/IEC/TR 80002-1:2009	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:2009	ANSI/AAMI/ISO 7199:2009	Identical
ISO 8637:2004	ANSI/AAMI RD16:2007	Major technical variations
ISO 8638:2004	ANSI/AAMI RD17:2007	Major technical variations
ISO 10993-1:2009	ANSI/AAMI/ISO 10993-1:2009	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003/(R)2009	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002/(R)2009 and Amendment 1:2006/(R)2009	Identical
ISO 10993-5:2009	ANSI/AAMI/ISO 10993-5:2009	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999/(R)2005	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002/(R)2008 ANSI/AAMI BE78:2002/A1:2006/(R)2008	Minor technical variations Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2009	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical

International designation	U.S. designation	Equivalency
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:2009	ANSI/AAMI/ISO 11737-2:2009	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003/(R)2009	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003/(R)2008	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003/(R)2008	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2009	ANSI/AAMI/ISO 14161:2009	Identical
ISO 14708-3:2008	ANSI/AAMI/ISO 14708-3:2008	Identical
ISO 14708-4:2008	ANSI/AAMI/ISO 14708-4:2008	Identical
ISO 14937:2009	ANSI/AAMI/ISO 14937:2009	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000/(R)2006 and A1:2004/(R)2006	Identical
ISO 15674:2009	ANSI/AAMI/ISO 15674:2009	Identical
ISO 15675:2009	ANSI/AAMI/ISO 15675:2009	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO 15883-1:2006	ANSI/AAMI ST15883-1:2009	Major technical variations
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical (with inclusions)
ISO/TS 17665-2:2009	ANSI/AAMI/ISO TIR17665-2:2009	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003/(R)2009 and A1:2005/(R)2009	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical
ISO 81060-2:2009	ANSI/AAMI/ISO 81060-2:2009	Identical

Committee representation

Association for the Advancement of Medical Instrumentation

Industrial Ethylene Oxide Sterilization Working Group

This technical information report (TIR) was developed by the AAMI Industrial Ethylene Oxide Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial Ethylene Oxide Sterilization Working Group** had the following members:

Cochairs Charles Cogdill, Boston Scientific Corporation
Gerry A. O'Dell, MS, Gerry O'Dell Consulting

Members Curt Andrew Bogue, BAS, Cook Inc.
Anne F. Booth, MS, Conmed Corporation
Lloyd Brown, Covidien
Bradley J. Catalone, PhD, Olympus America Inc.
Dennis E. Christensen, BS, Process Challenge Devices LLC
Charlie Christianson, St. Jude Medical Inc.
Charles Cogdill, Boston Scientific Corporation
Gary N. Cranston, Consulting & Technical Services/PCS
Elaine Daniell, CR Bard
Douglas D. Davie, Sterilization Validation Services
Shawn A. Doyle, Sterilator Company Inc.
William F. FitzGerald, PE, FitzGerald & Associates Ltd.
Dan B. Floyd, RM, Nelson Laboratories Inc.
Zory R. Glaser, PhD MPH CSPDM, Johns Hopkins University–School of Public Health (Independent Expert)
Arthur C. Harris, Cook Inc.
Deborah A. Havlik, Hospira Worldwide Inc.
Danny Hutson, Cardinal Health (MP&S)
Jim Kaiser, Bausch & Lomb Inc.
Bert Kingsbury, Terumo Medical Corporation
Carolyn L. Kinsley, LexaMed
Karen A. Kowalczyk, Centurion Sterilization Services
Christine Loshbaugh, Edwards LifeSciences
Ted May, Andersen Products Inc.
David Ford McGoldrick, BS, Abbott Laboratories
Craig A. Meadows, Medtronic Inc.
Joseph M. Mello, Ethide Laboratories Inc.
Russell D. Mills, Zimmer Inc.
Gary Mitchel, PE, Johnson & Johnson
Sarah A. Mowitt (Independent Expert)
Gerry A. O'Dell, MS, Gerry O'Dell Consulting
Ken Paddock, Baxter Healthcare Corporation
Dave Parente, NAMSA
Manuel Saavedra Jr., Kimberly-Clark Corporation
Zenius V. Seliokas, Stericon Inc.
Jon Seulean, CaridianBCT Sterilization Services Inc.
Barb Smith, Getinge USA
Bill South, Steris Corporation
Ralph Stick, WuXi AppTec
Wayne Swallow, Becton Dickinson & Company
Radhakrishna S. Tirumalai, US Pharmacopeial Convention Inc.
Steven E. Turtill, FDA/CDRH
Jason Voisinet, Ethox International Inc.

Craig A. Wallace, 3M Healthcare
P. Richard Warburton, ChemDAQ Inc.
Richard L. Weisman, Fresenius Medical Care Renal Therapies Group
Casimir John Woss, PhD, Alcon Laboratories Inc.
Stephen C. Yeadon, BS, Kinetic Concepts Inc.
William T. Young, Sterigenics International

Alternates

August Baur, Centurion Sterilization Services
John Broad, NAMSA
Delores Bruce, Steris Corporation
Trabue D. Bryans, WuXi AppTec
Susan Bullis, Johnson & Johnson
Tim Carlson, Becton Dickinson & Company
Kate Davenport, Baxter Healthcare Corporation
John DiCaro, Covidien
Brian R. Drumheller, CR Bard
Joseph R. Durbin, Hospira Worldwide Inc.
Mark Fischer, Nelson Laboratories Inc.
David Michael Gasparik, Cardinal Health (MP&S)
Thomas L. Hansen, Terumo Medical Corporation
Donna Horner, Abbott Laboratories
Brent Huberty, St. Jude Medical Inc.
Ezra Koski, A, CaridianBCT Sterilization Services Inc.
John M. Kuchinski, FDA/CDRH
James P. Kulla, BS MS, LexaMed
Richard Lenz, Medtronic Inc.
John Lindley, Andersen Products Inc.
Ralph Makinen, Boston Scientific Corporation
Jeff Martin, Alcon Laboratories Inc.
Janette Martinez, BS, Edwards LifeSciences
Michael E. Mattison, Getinge USA
Michael A. Padilla, BSME, Sterigenics International
Frank Peacock, Jr., Bausch & Lomb Inc.
Jeff Peltier, Boston Scientific Corporation
Nancy Rakiewicz, Ethox International Inc.
Tyron S. Rouse, Kimberly-Clark Corporation
David Silor, Zimmer Inc.
Larry Talapa, 3M Healthcare

AAMI also acknowledges the **TIR15 Editing Task Group**, comprising the following members, for its special contribution in the development of this document:

Task group leader Craig A. Meadows, Medtronic Inc.

Members Tim Carlson, Becton Dickinson & Company
Dennis E. Christensen, BS, Process Challenge Devices LLC
Jim Kaiser, Bausch & Lomb Inc.
Craig A. Meadows, Medtronic Inc.
Bill South, Steris Corporation

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

Cochairs Victoria M. Hitchins, PhD, FDA/CDRH
William E. Young, Boston Scientific Corporation

Members Trabue D. Bryans, WuXi AppTec
Peter A. Burke, PhD, Steris Corporation
Nancy Chobin, RN CSPDM, St. Barnabas Healthcare System (Independent Expert)
Charles Cogdill, Boston Scientific Corporation
Ramona Conner, RN MSN CNOR, Association of periOperative Registered Nurses
Jacqueline Daley, Association for Professionals in Infection Control and Epidemiology

Kimbrell Darnell, CR Bard
Lisa Foster, Sterigenics International
Joel R. Gorski, PhD, NAMSA
Deborah A. Havlik, Hospira Worldwide Inc.
Victoria M. Hitchins, PhD, FDA/CDRH
Danny Hutson, Cardinal Health (MP&S)
Lois Atkinson Jones, MS (Independent Expert)
Susan G. Klacik, CCSMC FCS ACE, IAHCSSM
Byron J. Lambert, PhD, Abbott Laboratories
Colleen Patricia Landers, RN, Canadian Standards Association
Lisa N. Macdonald, Becton Dickinson & Company
Jeff Martin, Alcon Laboratories Inc.
Patrick J. McCormick, PhD, Bausch & Lomb Inc.
Rainer Newman, Johnson & Johnson
Janet M. Prust, 3M Healthcare
Nancy Rakiewicz, Ethox International Inc.
Michael H. Scholla, Dupont Nonwovens
Mark Seybold, Baxter Healthcare Corporation
Andrew Sharavara, PhD, Propper Manufacturing Company Inc.
Mark N. Smith, Getinge USA
William N. Thompson, Covidien
James L. Whitby, MA MB FRCP (Independent Expert)
Martell Kress Winters, BS SM, Nelson Laboratories Inc.

Alternates

Lloyd Brown, Covidien
Glenn W. Calvert, Becton Dickinson & Company
Dave Dion, Cardinal Health (MP&S)
Steven J. Elliott, WuXi AppTec
Thomas J. Frazar, Johnson & Johnson
Kathy Hoffman, Sterigenics International
Jim Kaiser, Bausch & Lomb Inc.
Joseph J. Lasich, BS, Alcon Laboratories Inc.
Chiu S. Lin, PhD, FDA/CDRH
Natalie Lind, IAHCSSM
Ralph Makinen, Boston Scientific Corporation
Mary S. Mayo, CR Bard
David Ford McGoldrick, BS, Abbott Laboratories
Jerry R. Nelson, PhD, Nelson Laboratories Inc.
Karen Polkinghorne, Dupont Nonwovens
Mike Sadowski, Baxter Healthcare Corporation
John R. Scoville Jr., Steris Corporation
Jason Voisinet, Ethox International Inc.
Craig A. Wallace, 3M Healthcare
Valerie Welter, Hospira Worldwide Inc.
William E. Young, Boston Scientific Corporation

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

Foreword

This document is part of a series of technical information reports (TIRs) intended for use in conjunction with ANSI/AAMI/ISO 11135-1. The other reports in the series are

- AAMI TIR14:2009, *Contract sterilization using ethylene oxide*;
- AAMI TIR16:2009, *Microbiological aspects of ethylene oxide sterilization*;
- AAMI TIR28:2009, *Product adoption and process equivalence for ethylene oxide sterilization*; and
- ANSI/AAMI/ISO TIR11135-2:2008, *Sterilization of health care products—Ethylene oxide—Part 2: Guidance on the application of ANSI/AAMI/ISO 11135-1*

The original TIR15, along with other AAMI TIRs, provided additional guidance to the 1994 edition of the industrial ethylene oxide sterilization standard ANSI/AAMI/ISO 11135, which was revised in 2007 under a new designation, ANSI/AAMI/ISO 11135-1:2007, *Sterilization of health care products—Ethylene oxide—Part 1: Requirements for development, validation, and routine control of a sterilization process for medical devices*. In 2008, the International Organization for Standardization (ISO) published its own guidance document for the 11135 standard, ISO/TR 11135-2:2008, *Sterilization of health care products—Ethylene oxide—Part 2: Guidance on the application of ISO 11135-1*, which was based to a great extent on the earlier AAMI TIRs. Correspondingly, the AAMI Industrial Ethylene Oxide Sterilization Working Group is updating its TIRs to take into account changes to the 11135 standard and to avoid redundancy with ANSI/AAMI/ISO TIR11135-2:2008.

This TIR contains guidelines that are not intended to be absolute or to apply in all circumstances. One should use judgment in applying the information in this TIR.

As used within the context of this document, “should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the technical information report. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation. See also the NOTE on Page 1.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

NOTE—This foreword does not contain provisions of AAMI TIR15:2009, *Physical aspects of ethylene oxide sterilization*, but it does provide important information about the development and intended use of the document.

Physical aspects of ethylene oxide sterilization

NOTE—This technical information report (TIR) is not a standard, and the material contained herein is not normative in nature. In a few instances, the committee has used the term “shall” based on its knowledge of requirements contained in relevant standards, regulatory requirements, or both.

1 Scope

This AAMI TIR provides additional guidance to augment ANSI/AAMI/ISO 11135-1:2007. This TIR addresses some of the physical aspects of ethylene oxide (EO) sterilization and provides guidance on sterilization equipment, considerations for preconditioning, how to calculate relative humidity, EO concentration, and flammability, as well as guidance on the use of statistics for process equivalence.

Additional guidance on ANSI/AAMI/ISO 11135-1:2007 can be found in ANSI/AAMI/ISO 11135-2:2008.

2 Definitions

For the purposes of this TIR, the terms and definitions in ANSI/AAMI/ISO 11135-1:2007, ANSI/AAMI/ISO 11135-2:2008, and the following apply:

- 2.1 dew point:** The temperature at which the saturation water vapor pressure is equal to the partial pressure of the water vapor in the atmosphere. Any cooling of the atmosphere below the dew point would produce water condensation.
- 2.2 humidity:** A measure of the water vapor present in a gas. It is usually measured as absolute humidity, relative humidity, or dew point temperature.
- 2.3 preconditioning area:** Either a chamber or a room in which product preconditioning occurs.
- 2.4 relative humidity (RH):** Ratio of water vapor partial pressure actually present to water vapor pressure required for saturation at a given temperature.

$$\%RH = \frac{\text{actual vapor pressure}}{\text{saturation vapor pressure}} \times 100$$

3 EO sterilization processing equipment

3.1 General

3.1.1 When 100 % EO or flammable blends of EO are used,

- a) Electrical accessories used for EO processes should comply with Class I, Division 2, Group B electrical requirements as stated by the National Fire Protection Association (NFPA) in NFPA 70 (2008) or equivalent. The guidance within NFPA 560 (2007) should also be consulted.
- b) The sterilizer interior should comply with Class I, Division 1, Group B electrical requirements stated within NFPA 70 (2008) or equivalent. The guidance within NFPA 560 (2007) should also be consulted.
- c) Equipment and piping should be grounded in accordance with NFPA 70 (2008) or equivalent.
- d) In facilities constructed after 1995 and where NFPA standards are under the jurisdiction of a federal, state, local, or other regional department or individual (e.g., fire chief), the storage, handling, and use of EO shall comply with NFPA 560 (2007).
- e) When designing sterilization cycles, the chamber environment should remain nonflammable.