

# Technical Information Report

## AAMI TIR56: 2013/(R)2016

Guidance for the  
development, validation  
and routine control of an  
ethylene oxide sterilization  
process utilizing flexible  
bag systems for the  
sterilization of medical  
devices

# Guidance for the development, validation and routine control of an ethylene oxide sterilization process utilizing flexible bag systems for the sterilization of medical devices

Approved 27 December 2013 and reaffirmed 18 October 2016 by  
AAMI

**Abstract:** This AAMI Technical Information Report (TIR) provides information to be considered during the development, validation, and routine control of EO sterilization processes that are performed using gas diffusion within individually sealed flexible sterilization bags.

**Keywords:** sterilization, flexible bag systems, diffusion, EO, ethylene oxide

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and in the case of American National Standards by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

*Published by*

AAMI  
4301 N. Fairfax Drive, Suite 301  
Arlington, VA 22203-1633  
[www.aami.org](http://www.aami.org)

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at [www.aami.org](http://www.aami.org) or contact AAMI at 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: +1-703-525-4890; Fax: +1-703-276-0793.

Printed in the United States of America

**ISBN 978-1-57020-519-4**

# Contents

Page

Glossary of equivalent standards.....	v
Committee representation.....	vi
Foreword .....	x
Introduction.....	xi
1 Scope .....	1
2 Terms and definitions.....	1
3 Process descriptions .....	3
3.1 Type 1 - Gas injection systems.....	3
3.2 Type 2 – EO cartridge based systems utilizing permeable flexible sterilization bags .....	4
3.3 Type 3 - EO cartridge based systems utilizing effectively impermeable flexible sterilization bags.....	5
4 Quality management systems.....	6
5 Sterilizing agent characterization .....	6
5.1 General.....	6
5.2 Sterilizing agent.....	6
5.3 Microbicidal effectiveness .....	6
5.4 Material effects.....	6
5.5 Safety and the environment .....	6
6 Process and Equipment Characterization.....	7
6.1 Process characterization.....	7
6.2 Equipment characterization .....	8
7 Product definition .....	9
8 Process definition.....	12
9 Validation .....	13
9.1 Installation qualification .....	13
9.2 Operational qualification.....	13
9.3 Performance qualification.....	15
9.4 Varying load configurations.....	17
9.5 Review and approval of validation .....	17
10 Routine Monitoring and Control .....	19
11 Product Release from Sterilization.....	20
12 Maintaining Process Effectiveness .....	20
12.1 General.....	20
12.2 Calibration .....	20
12.3 Maintenance of equipment.....	20
12.4 Requalification.....	21
12.5 Assessment of change.....	21
12.6 Assessment of equivalence .....	22

**Annexes**

A Microbial Validation Using an Augmented Overkill Approach ..... 24

B Microbial Validation Using Reduced EO Concentration ..... 26

Bibliography ..... 27

**Tables and Figures**

Table 1 — Flexible bag sterilization system types ..... 3

Figure 1 — Sample PCD styles (A & B) ..... 23

Figure 2 — How the overkill or SAL of  $10^{-6}$  is achieved using the internal PCD (BI in EO cartridge) ..... 24

Figure 3 — Comparison of theoretical EO concentration (lethality) profiles for full production cycle and validation cycle using half the production EO concentration. Validation lethality is equal to  $\frac{1}{2}$  the exposure and  $\frac{1}{4}$  the lethality obtained during EO injection and removal, based on using the same injection and evacuation rates ..... 25

## **Glossary of equivalent standards**

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

[www.aami.org/standards/glossary.pdf](http://www.aami.org/standards/glossary.pdf)

## Committee representation

### Association for the Advancement of Medical Instrumentation

#### Industrial EO Sterilization Working Group

This Technical Information Report (TIR) was developed by the AAMI Industrial EO Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial EO Sterilization Working Group** had the following members:

**Chairs:** Jeff Martin, Alcon Laboratories Inc  
Gerry A. O'Dell, MS, Gerry O'Dell Consulting

**Members:** Edward Arscott, NAMSA  
Anne F. Booth, MS, Booth Scientific Inc  
Lloyd Brown, Covidien  
Tim Carlson, Becton Dickinson & Company  
Dennis E. Christensen, BS, Process Challenge Devices LLC  
Gary N. Cranston, Consulting & Technical Services/PCS  
Elaine Daniell, CR Bard  
Douglas D. Davie, Sterilization Validation Services  
Darci Diage, Direct Flow Medical Inc  
Mary Ann Drosnock, MS, Olympus America Inc  
William F. FitzGerald, PE, FitzGerald & Associates Ltd  
Dan B. Floyd, Nelson Laboratories Inc  
Sarah Gagnon, Microtest Laboratories Inc  
Naomi Gamm, St Jude Medical Inc  
Zory R. Glaser, PhD MPH CSPDM, Johns Hopkins University-School of Public Health  
Michael Groendyk, Javelin Medical  
Thomas Hansen, Terumo Americas Corporate  
Douglas F. Harbrecht, Sterility Assurance LLC  
Arthur C. Harris, Cook Inc  
Deborah A. Havlik, Hospira Worldwide Inc  
Jason Hedrick, Medtronic Inc WHQ Campus  
Danny Hutson, CareFusion  
Nupur Jain, Intuitive Surgical Inc  
Jim Kaiser, Bausch & Lomb Inc  
David King, Tandem Diabetes Care Inc  
Carolyn L. Kinsley, LexaMed Ltd  
Mindy Kobus, Mesa Laboratories Biological Indicator Division- Bozeman Facility  
Karen A. Kowalczyk, Centurion Sterilization Services  
Reynaldo Lopez, Cardinal Health (MP&S)  
Christine Loshbaugh, Edwards LifeSciences  
Ralph Makinen, Boston Scientific Corporation  
Jeff Martin, Alcon Laboratories Inc  
Albert E. May, Andersen Products Inc  
David Ford McGoldrick, BS, Abbott Laboratories  
Russell D. Mills, GE Healthcare  
Gary Mitchel, PE, Johnson & Johnson  
Sarah A. Mowitt  
Dave Parente, Ecolab  
Michelle Peterson, Stryker Instruments Division

Nancy Rakiewicz, Moog Medical Devices  
Manuel Saavedra, Jr., Kimberly-Clark Corporation  
Steven Seamons, WL Gore & Associates Inc  
Jon Seulean, Terumo BCT  
Mark Seybold, Baxter Healthcare Corporation  
Harry L. Shaffer, Sterilization Consulting Services  
David Silor, Zimmer Inc  
Bill South, Steris Corporation  
Michael Sprague, Ethide Laboratories Inc  
Sopheak Srun, MPH SM(NRCM), Quality Tech Services Inc  
Ralph Stick, WuXi AppTec Inc  
Fenil Sutaria, Medline Industries Inc  
Radhakrishna S. Tirumalai, US Pharmacopeia Convention Inc  
Steven E. Turtill, FDA/CDRH  
Craig A. Wallace, 3M Healthcare  
P. Richard Warburton, ChemDAQ Inc  
Richard L. Weisman, Fresenius Medical Care Renal Therapies Group  
Beverly G. Whitaker, CQA RAB MBA, Indigo Consulting Group Inc  
William T. Young, Sterigenics International

**Alternates:**

David E. Barlow, PhD, Olympus America Inc  
August Baur, Centurion Sterilization Services  
Marjean Boyter, Fresenius Medical Care Renal Therapies Group  
Robert H. Bradley, Mesa Laboratories Biological Indicator Division-Bozeman Facility  
Carolyn Braithwaite-Nelson, Spectranetics Corporation  
Trabue D. Bryans, BryKor LLC  
Susan Bullis, Johnson & Johnson  
Greg Bush, Alcon Laboratories Inc  
Claudia Camp, Stryker Instruments Division  
Phil Cogdill, Covidien  
Greg Crego, Moog Medical Devices  
Jessica Desmond, Microtest Laboratories Inc  
John DiCaro, Covidien  
Dave Dion, Cardinal Health (MP&S)  
David A. Dominguez, CareFusion  
Brian R. Drumheller, CR Bard  
Steven Elliott, FDA/CDRH  
Scott A. Giraud, Medtronic Inc WHQ Campus  
William K. Gordon, Steris Corporation  
Matthew Hart, Boston Scientific Corporation  
Sharon Higgins, GE Healthcare  
David M. Hilliker, ChemDAQ Inc  
Ezra Koski, A, Terumo BCT  
James P. Kulla, BS MS, LexaMed Ltd  
Wesley Lantz, Zimmer Inc  
John Lindley, Andersen Products Inc  
Paul Littley, Nelson Laboratories Inc  
Ronald G. Lulich, 3M Healthcare  
Jody L. Markling, St Jude Medical Inc  
Janette Martinez, BS, Edwards Lifesciences (PR)  
Patrick J. McCormick, PhD, Bausch & Lomb Inc  
Joseph Mello, Ethide Laboratories Inc  
Laurie Nawrocki, NAMSA  
Koyejo Obadina, Abbott Laboratories  
Mike Padilla, Sterigenics International  
Andrew Porteous, Baxter Healthcare Corporation

Tyrone S. Rouse, Kimberly-Clark Corporation  
Matthew Russell, Cook Inc  
Arnold Shechtman, BS, Medline Industries Inc  
Brian Wallace, Intuitive Surgical Inc  
Daryl Woodman, Andersen Products Inc  
Casimir John Woss, PhD, FitzGerald & Associates Ltd

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

**Chairs:** Victoria M. Hitchins, PhD, FDA/CDRH  
Michael H. Scholla, PhD, Dupont Protection Technologies

**Members:** Christopher Anderson, Boston Scientific Corporation  
Trabue D. Bryans, BryKor LLC  
Nancy Chobin, RN CSPDM, St Barnabas Healthcare System  
Charles Cogdill, Covidien  
Ramona Conner, RN MSN CNOR, Association of Perioperative Registered Nurses  
Jacqueline Daley, Sinai Hospital of Baltimore  
Kimbrell Darnell, CR Bard  
Lisa Foster, Medpoint LLC  
Joel R. Gorski, PhD, NAMSA  
Joyce M. Hansen, Johnson & Johnson  
Douglas F. Harbrecht, Sterility Assurance LLC  
Deborah A. Havlik, Hospira Worldwide Inc  
Susan G. Klacik, CCSMC FCS ACE, IAHCMM  
Byron J. Lambert, PhD, Abbott Laboratories  
Colleen Patricia Landers, RN, Timmins & District Hospital  
Reynaldo Lopez, Cardinal Health (MP&S)  
Jeff Martin, Alcon Laboratories Inc  
Patrick J. McCormick, PhD, Bausch & Lomb Inc  
Gerald E. McDonnell, PhD, Steris Corporation  
Janet M. Prust, 3M Healthcare  
Nancy Rakiewicz, Moog Medical Devices  
Mark Seybold, Baxter Healthcare Corporation  
Andrew Sharavara, PhD, Propper Manufacturing Co Inc  
Mark N. Smith, Getinge USA  
James Sidney Wiggs, BSN CRCST, Legacy Health System  
Martell Kress Winters, BS SM, Nelson Laboratories Inc  
William T. Young, Sterigenics International

**Alternates:** Lloyd Brown, Covidien  
Peter A. Burke, PhD, Steris Corporation  
Dave Dion, Cardinal Health (MP&S)  
Ken Eddington, NAMSA  
Gordon M. Ely, WuXi AppTec Inc  
Thomas J. Frazar, Johnson & Johnson  
Martha M. Kadas, Sterigenics International  
Jim Kaiser, Bausch & Lomb Inc  
Natalie Lind, IAHCMM  
Ralph Makinen, Boston Scientific Corporation  
Mary S. Mayo, CR Bard  
David Ford McGoldrick, BS, Abbott Laboratories

Jerry R. Nelson, PhD, Nelson Laboratories Inc  
Patrick Polito, Moog Medical Devices  
Karen Polkinghorne, Dupont Protection Technologies  
Mike Sadowski, Baxter Healthcare Corporation  
Craig A. Wallace, 3M Healthcare

---

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

The processes, methods, and equipment described in this TIR vary significantly from those described in ANSI/AAMI/ISO 11135-1:2007. These differences include but are not limited to the following:

- The body of information provided in ANSI/AAMI/ISO 11135-1:2007 is complemented by supporting documentation found in ANSI/AAMI/ISO 11135-2:2008, as well as a number of related AAMI TIR documents, significant parts of which may not apply to methods and materials that are the subject of this TIR.
- The 11135 series includes 100 % verification of chamber installation and the associated performance qualification (physical and microbiological) of fixed equipment; the methods described in this document rely upon the validation of an integrated, flexible bag/chamber design and a high degree of quality assurance during the manufacturing process of the flexible bag/chamber (and other accessories).
- Methods included in this document identify processes based solely on gas injection by weight, variable, or diminishing (i.e., decreasing by diffusion through the flexible bag) ethylene oxide (EO) concentrations in the atmosphere external to the package but internal to the “flexible bag sterilization systems” per the design of the method.
- Approaches to validation, routine monitoring, and control of cycles may differ significantly from those indicated in 11135.
- For some process designs, the lethality of the process relies completely on storage conditions and time (post introduction of EO) to effectively provide the requisite sterility assurance level. In addition, the lethality provided during the storage period within the sterilization/aeration cabinets may also act simultaneously in providing the aeration method. There are also a number of specialized and novel equipment designs necessary to remove EO from a flexible sterilization bag that vary considerably from the methods described in 11135. New considerations may arise regarding the identification of worst-case aeration processes, as well as subsequent validation.

## Introduction

As a result of a revision to ANSI/AAMI/ISO 11135-1, the exclusion in the document stating “This part of ISO 11135 does not cover sterilization by injecting ethylene oxide or mixtures containing ethylene oxide directly into individual product packages,” was revised to state, “This International Standard does not cover sterilization by injecting EO or mixtures containing EO directly into packages **or a flexible chamber.**” The addition of the flexible chamber to the international standard clarified the intent that ANSI/AAMI/ISO 11135 should only be applied to traditional fixed sterilization chambers, and that ANSI/AAMI/ISO 14937:2009/(R)2013 should be used for the requirements and guidance related to these types of EO processes. Ethylene oxide sterilization is a long-standing sterilization process, and EO sterilization within flexible sterilization bags has been used commercially since the 1960s. However, little published information was available to provide guidance on adequately addressing development, validation, and routine control of these EO systems. Therefore, the Industrial EO Sterilization Working Group chose to develop a technical information report (TIR) that addresses these aspects as a prelude to the eventual development of an international standard.

During the development of the TIR, there was strong objection by some to the use of EO sterilization within a flexible sterilization bag; therefore, this document does *not* represent a committee or national consensus. Most of this objection centered around two specific areas of concern. The first objection was the lack of workplace safety information that is provided in this TIR. The decision not to include safety related details within this TIR is consistent with the position that has been taken in other industrial sterilization standards or guidance documents. Information related to the safe use of EO in the workplace is readily available (see <https://www.osha.gov/SLTC/ethyleneoxide/>) and its use falls under the regulatory auspices of OSHA (29 CFR 1910.1047) (2005). Additional information related to the carcinogenic properties of EO can be found in World Health Organization, 2008.

The second objection was related to the dearth of published information related to EO sterilization in flexible sterilization bags that documented its efficacy or repeatability. The biggest differences between traditional EO sterilization and this non-traditional method are the use of the flexible sterilization bag with a very low concentration of EO and the absence of significant air removal during the process. While there is not significant literature related to this process in particular, there is considerable literature on the efficacy and repeatability of EO sterilization in general, even under conditions that are somewhat similar to those used during EO sterilization in a flexible sterilization bag ([1] and [2]). During the development of this TIR, two articles were published ([5] [6]) that have particular relevance to this process. The expectation is that as the TIR is used, additional information will be made public to further substantiate its efficacy and repeatability, as well as to refine how the processes are developed, validated and routinely controlled.

# Guidance for the development, validation, and routine control of an ethylene oxide sterilization process utilizing flexible bag systems for the sterilization of medical devices

## 1 Scope

This TIR includes information to be considered during the development, validation, and routine control of EO sterilization processes that are performed using gas diffusion within individually sealed flexible sterilization bags.

## 2 Terms and definitions

### 2.1

#### **abatement system**

device that is attached to the exhaust from the sterilization/aeration cabinet and connected to the outside atmosphere. It is designed to absorb and/or react with the EO gas exhausted from the cabinet and reduce the amount of EO exhausted to the external atmosphere

### 2.2

#### **aeration cycle**

part of the sterilization process, after the sterilization portion of the cycle, during which EO and/or its reaction products dissipate from the device(s)

### 2.3

#### **chemical integrator**

demonstrates that the sterilization parameters over a specified range of sterilization cycles have been met in a specified sterilization wrap, container, cassette, or pouch

### 2.4

#### **dunnage**

material that duplicates the weight, volume, thermal characteristics, EO absorption characteristics, surface geometry and tortuosity, and other critical properties of the devices being tested

### 2.5

#### **effectively impermeable**

term used to describe a bag material that does not allow for diffusion of EO during the sterilization cycle

### 2.6

#### **EO cartridge**

single use, hermetically sealed container that holds a predetermined weight of EO. The EO cartridge is designed to be manually activated through the flexible material of the flexible sterilization bag, releasing 100 % EO

### 2.7

#### **exposure time**

period for which the process parameters (temperature, relative humidity, and EO concentration ) are maintained within specified tolerances