

# Technical Information Report

## AAMI TIR63: 2014

Management of loaned  
critical and semi-critical  
medical devices that  
require sterilization or  
high-level disinfection



# Management of loaned critical and semi-critical medical devices that require sterilization or high-level disinfection

Approved 23 December 2014 by  
Association for the Advancement of Medical Instrumentation

**Abstract:** This technical information report identifies the necessary steps to effectively manage medical devices not owned by the health care facility in which they are used.

**Keywords:** high-level disinfection, loaned medical device, medical equipment management, quality systems, sterilization

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are voluntary, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633

*Published by*

Association for the Advancement of Medical Instrumentation  
4301 N. Fairfax Dr., Ste. 301  
Arlington, VA 22203-1633  
[www.aami.org](http://www.aami.org)

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 4301 N. Fairfax Dr., Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

**ISBN 1-57020-574-4**

**Contents**

Page

Glossary of equivalent standards ..... iv

Committee representation ..... v

Foreword ..... viii

Introduction ..... 1

1 Scope ..... 1

    1.1 General ..... 1

    1.2 Inclusions ..... 1

    1.3 Exclusions ..... 1

2 Definitions and abbreviations ..... 1

3 Policy development ..... 2

4 Procedures ..... 2

    4.1 General considerations ..... 2

    4.2 Roles and responsibilities of the sender ..... 2

    4.3 Roles and responsibilities of the receiver ..... 3

        4.3.1 Communication ..... 3

        4.3.2 Financial responsibility ..... 3

        4.3.3 Device responsibility ..... 3

    4.4 Requisition ..... 3

    4.5 Scheduling ..... 4

        4.5.1 Delivery ..... 4

        4.5.2 Information from sender ..... 4

        4.5.3 Processing time ..... 4

    4.6 Receipt ..... 4

    4.7 Return ..... 4

    4.8 Tracking and documentation ..... 5

        4.8.1 General considerations ..... 5

        4.8.2 Documentation ..... 5

        4.8.3 Receipt documentation ..... 5

Bibliography ..... 6

## **Glossary of equivalent standards**

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

[www.aami.org/standards/glossary.pdf](http://www.aami.org/standards/glossary.pdf)

## Committee representation

### Association for the Advancement of Medical Instrumentation

#### Reusable Sterilization Container Working Group

This standard was developed by the AAMI Reusable Sterilization Container Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the standard does not necessarily mean that all working group members voted for its approval. At the time this standard was published, the **AAMI Reusable Sterilization Container Working Group** had the following members:

*Chairs:* Damien Berg, Medical Center of the Rockies  
Joan M. Spear, MBA RM CMPR, B Braun of America Inc

*Members:* Mohamed Ali, Medline Industries Inc  
Heide M. Ames, Steris Corporation  
Angela H. Brightwell, Medtronic Inc WHQ Campus  
Renee Camp, MS, Moog Medical Devices  
Jason Cartwright, CareFusion  
Nancy Chobin, RN CSPDM, St Barnabas Healthcare System  
Susan Christensen, Christensen Consulting  
Linda Condon, Johns Hopkins Hospital  
Ramona Conner, RN MSN CNOR, Association of Perioperative Registered Nurses  
Jacqueline Daley, Sinai Hospital of Baltimore  
April J. Doering, St Jude Medical Inc  
Mary Ann Drosnock, MS, Olympus America Inc  
Betty D. Edge, North Shore University Hospital  
Steven J. Elliott, NAMSA  
Steven Elliott, FDA/CDRH  
Sue Ellen Erickson, MS CNOR, Newark Beth Israel Medical Center  
Susan Flynn, 3M Healthcare  
Marcia Ann Frieze, Case Medical Inc  
Dixit Gandhi, Medline Industries Inc  
Shelley Green, WuXi AppTec Inc  
Charles Oren Hancock, RAC, H&W Technology LLC  
Jason Harrington, VA Medical Center of Cincinnati  
Sybil Hickes, Materials Management Microsystems  
Charles A. Hughes, Medivators Inc  
Nupur Jain, Intuitive Surgical Inc  
Scott A. Jelley, Johnson & Johnson  
David W. Johnson, Kimberly-Clark Corporation  
Susan G. Klacik, CCSMC FCS ACE, IAHCSSM  
Colleen Patricia Landers, RN, Timmins & District Hospital  
Mary Kneece Lane, BS MHA CSPDS CSPDM  
Helene Leblond, TSO3 Inc  
Teckla A. Maresca, LPN CSPDM, St Clare's Health System  
Emily Mitzel, MS, Nelson Laboratories Inc  
Thomas K. Moore, C-Moore  
Frank Myers, USPHS Indian Health Service  
Karen Nauss, CRCST, Mount Auburn Hospital  
Gerry A. O'Dell, MS, Gerry O'Dell Consulting  
Rodney D. Parker, Stryker Instruments Division  
Anthony Powell, Getinge USA  
Rose E. Seavey, RN MBA CNOR CRCST, Seavey Healthcare Consulting, LLC  
Frank Sizemore, Wake Forest University Baptist Medical Center  
Linda Slone, RN BSPA CNOR  
Gary J. Socola, HIGHPOWER Validation Testing & Lab Services Inc  
Betty Strickland, Pryce Consultants  
Karen Swanson, Connecticut Childrens Medical Center  
R. Brent Sweet, Zimmer Inc  
Raymond Taurasi, MBA CRCST CHL FCS AC, Healthmark Industries Company Inc  
Nora E. Wikander, RN, CSPDM, St Josephs Wayne Hospital  
Ann Young, Fletcher Allen Health Care  
Martha L. Young, Martha L Young, LLC

*Alternates:* Terence A. Alwin, Medline Industries Inc  
Ralph J. Basile, MBA, Healthmark Industries Company Inc  
Stacy DeMoss, NAMSA  
Sylvie Dufresne, PhD, TSO3 Inc  
Gordon M. Ely, WuXi AppTec Inc  
Jeff Felgar, Zimmer Inc  
Naomi Gamm, St Jude Medical Inc  
Rachel Hill, CareFusion  
Mariann Hughes, Medivators Inc  
Natalie Lind, IAHCMM  
Tania Lupu, Case Medical Inc  
Michael Neilson, Nelson Laboratories Inc  
Edward Nuber, B Braun of America Inc  
Anthony M. Piotrkowski, Steris Corporation  
Patrick Polito, Moog Medical Devices  
Janet M. Prust, 3M Healthcare  
Mandy Ryan, Stryker Instruments Division  
Barb Smith, Getinge USA  
Shani Smith, OMPT/CDRH/ODE/DAGID/INCB  
Donald Tumminelli, HIGHPOWER Validation Testing & Lab Services Inc  
Donna Ungvarsky, MSN MEd RN CNOR, Olympus America Inc  
Sharon Van Wicklin, MSN RN CNOR/CRNFA, Association of Perioperative Registered Nurses

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

*Co-chairs:* Victoria M. Hitchins, PhD, FDA/CDRH  
Michael H. Scholla, Dupont Protection Technologies

*Members:* Christopher Anderson, Boston Scientific Corporation  
Trabue D. Bryans, BryKor LLC  
Nancy Chobin, RN CSPDM, St Barnabas Healthcare System  
Charles Cogdill, Covidien  
Ramona Conner, RN MSN CNOR, Association of Perioperative Registered Nurses  
Jacqueline Daley, Sinai Hospital of Baltimore  
Kimbrell Darnell, CR Bard  
Lisa Foster, Medpoint LLC  
Joel R. Gorski, PhD, NAMSA  
Joyce M. Hansen, Johnson & Johnson  
Douglas F. Harbrecht, Sterility Assurance LLC  
Deborah A. Havlik, Hospira Worldwide Inc  
Susan G. Klacik, CCSMC FCS ACE, IAHCMM  
Byron J. Lambert, PhD, Abbott Laboratories  
Colleen Patricia Landers, RN, Timmins & District Hospital  
Reynaldo Lopez, Cardinal Health (MP&S)  
Jeff Martin, Alcon Laboratories Inc  
Patrick J. McCormick, PhD, Bausch & Lomb Inc  
Gerald E. McDonnell, PhD, Steris Corporation  
Janet M. Prust, 3M Healthcare  
Nancy Rakiewicz, Moog Medical Devices  
Mark Seybold, Baxter Healthcare Corporation  
Andrew Sharavara, PhD, Propper Manufacturing Co Inc  
Mark N. Smith, Getinge USA  
James Sidney Wiggs, BSN CRCST, Legacy Health System  
Martell Kress Winters, BS SM, Nelson Laboratories Inc  
William E. Young  
William T. Young, Sterigenics International

*Alternates:* Lloyd Brown, Covidien  
Peter A. Burke, PhD, Steris Corporation  
Dave Dion, Cardinal Health (MP&S)  
Gordon M. Ely, WuXi AppTec Inc  
Thomas J. Frazar, Johnson & Johnson  
Martha M. Kadas, Sterigenics International  
Jim Kaiser, Bausch & Lomb Inc  
Natalie Lind, IAHCMM

Ralph Makinen, Boston Scientific Corporation  
Mary S. Mayo, CR Bard  
David Ford McGoldrick, BS, Abbott Laboratories  
Jerry R. Nelson, PhD, Nelson Laboratories Inc  
Patrick Polito, Moog Medical Devices  
Karen Polkinghorne, Dupont Protection Technologies  
Shaundrea L. Rechsteiner, NAMSA  
Mike Sadowski, Baxter Healthcare Corporation  
Craig A. Wallace, 3M Healthcare

---

NOTE—Participation by federal agency representatives in the development of this recommended practice does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

This technical information report (TIR) was developed by the AAMI Rigid Sterilization Container System Working Group under the auspices of the AAMI Sterilization Standards Committee.

Management of a loaned medical devices presents challenges to both the sending and receiving facilities. Currently, there is much variation in management of loaned medical devices across health care facilities. This TIR was developed to provide a consistent approach to management of loaned medical devices by defining responsibilities and addressing critical requirements such as ensuring there is sufficient time to process a loaned device after delivery and before use.

As used within the context of this document, “should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the technical information report. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

Suggestions for improving this recommended practice are invited. Comments and suggested revisions should be sent to Technical Programs, AAMI, 4301 N. Fairfax Dr., Ste. 301, Arlington, VA 22203-1633.

# Management of loaned critical and semi-critical medical devices that require sterilization or high-level disinfection

## Introduction

Health care facilities frequently need to obtain medical devices they do not own in order to provide for an individual patient need. The facility may obtain these devices through a loan agreement with a vendor or another health care facility. The devices should be delivered to the receiving facility in a timely manner, cleaned and sterilized or high-level disinfected by the receiving health care facility or other designated facility prior to patient use, delivered to the point of care on schedule, cleaned and rendered safe to handle following use, and returned to the sender in good condition. Effective and safe management of loaned medical devices is a complex endeavor requiring systematic processes and clear communication among all involved parties.

This technical information report (TIR) is intended to provide guidance for the management of loaned medical devices. Only medical devices that are cleaned and sterilized or high-level disinfected according to the manufacturer's written IFU and functioning according to manufacturer's specifications should be used in a medical procedure. A formal program between the health care facility and the lending entity is necessary to ensure that loaned devices are properly handled, readied for patient use, and returned undamaged.

## 1 Scope

### 1.1 General

This TIR identifies the necessary steps to effectively manage medical devices not owned by the health care facility in which they are used.

### 1.2 Inclusions

This TIR addresses the management of loaned critical and semi-critical medical devices that require sterilization or high-level disinfection by the receiving facility prior to patient use. This document includes processes for the request, receipt, return, and documentation of loaned medical devices. Included are the roles and responsibilities of the health care facility ("receiver"), the lending entity ("sender"), and individuals.

### 1.3 Exclusions

This TIR does not address department design considerations, personnel considerations, cleaning and other decontamination processes, packaging, preparation, or sterilization (see ANSI/AAMI ST79, *Comprehensive guide to steam sterilization and sterility assurance in health care facilities*; ANSI/AAMI ST41, *Ethylene oxide sterilization in health care facilities: Safety and effectiveness*; and ANSI/AAMI ST58, *Chemical sterilization and high-level disinfection in health care facilities*). Processes for management of non-critical loaned medical devices and single-use medical devices, as well as off-site reprocessing facilities, are outside the scope of this document.

## 2 Definitions and abbreviations

**2.1 health care facility:** hospital, free-standing surgical center, clinic, medical office, or dental office.

**2.2 loaned:** A medical device intended for use by a health care facility that is not owned by the facility.

**2.3 medical device:** Instrument, apparatus, material, or other article, whether used alone or in combination, including the software necessary for its proper application, intended by the manufacturer to be used for human beings for the purpose of

- diagnosis, prevention, monitoring, treatment, or alleviation of disease;
- diagnosis, monitoring, treatment, alleviation of, or compensation for an injury or handicap;
- investigation, replacement, or modification of the anatomy or of a physiological process; or
- control of conception