

CONSOLIDATED VERSION

VERSION CONSOLIDÉE



**Medical electrical equipment – Medical image display systems –
Part 1: Evaluation methods**

**Appareils électromédicaux – Systèmes d'imagerie médicale –
Partie 1: Méthodes d'évaluation**



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2016 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 15 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - www.iec.ch/searchpub

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 15 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.



IEC 62563-1

Edition 1.1 2016-03

**CONSOLIDATED
VERSION**

**VERSION
CONSOLIDÉE**



**Medical electrical equipment – Medical image display systems –
Part 1: Evaluation methods**

**Appareils électromédicaux – Systèmes d'imagerie médicale –
Partie 1: Méthodes d'évaluation**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.55

ISBN 978-2-8322-3270-5

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

REDLINE VERSION

VERSION REDLINE



**Medical electrical equipment – Medical image display systems –
Part 1: Evaluation methods**

**Appareils électromédicaux – Systèmes d'imagerie médicale –
Partie 1: Méthodes d'évaluation**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
INTRODUCTION to Amendment 1	7
1 Scope	8
2 Normative references	8
3 Terms, definitions, symbols and abbreviations.....	8
3.1 Terms and definitions	8
3.2 Symbols	11
3.3 Abbreviations	12
4 General.....	12
5 Prerequisites	12
6 Equipment and tools	13
6.1 LUMINANCE meter	13
6.2 ILLUMINANCE meter.....	13
6.3 Colour meter	13
6.4 TEST PATTERNS	14
7 Evaluation methods	15
7.1 General.....	15
7.2 Evaluation method table overview	15
7.3 Visual evaluation methods	17
7.3.1 General	17
7.3.2 Overall image quality evaluation.....	17
7.3.3 Greyscale resolution evaluation.....	18
7.3.4 LUMINANCE response evaluation.....	19
7.3.5 LUMINANCE uniformity evaluation.....	20
7.3.6 Chromaticity evaluation	20
7.3.7 Pixel faults evaluation	20
7.3.8 VEILING GLARE evaluation.....	21
7.3.9 Geometrical image evaluation	21
7.3.10 Angular viewing evaluation.....	22
7.3.11 Clinical evaluation.....	23
7.4 Quantitative evaluation methods	23
7.4.1 Basic LUMINANCE evaluation.....	23
7.4.2 Basic LUMINANCE evaluation without ambient light	24
7.4.3 LUMINANCE response evaluation.....	24
7.4.4 LUMINANCE evaluation of multiple displays	27
7.4.5 Chromaticity uniformity evaluation	27
7.4.6 Chromaticity evaluation of across multiple displays.....	27
7.4.7 LUMINANCE uniformity evaluation.....	27
7.4.8 Viewing angle evaluation.....	28
7.4.9 Greyscale chromaticity evaluation	28
Annex A (informative) Sample test reports.....	29
Annex B (informative) LUMINANCE measurement methods	48
Annex C (informative) Description of TEST PATTERNS	51
Bibliography	60

Index of defined terms	62
Figure 1 – Overall image quality evaluation using the TG18-QC TEST PATTERN	17
Figure 2 – Overall image quality evaluation using the TG18-OIQ TEST PATTERN	18
Figure 3 – Magnified view of TG18-MP TEST PATTERN showing the 8-bit and 10-bit markers	19
Figure 4 – A close-up of the TG18-CT TEST PATTERN.....	20
Figure 5 – The TG18-GV TEST PATTERN is displayed (left), a close-up of the centre of the TEST PATTERN when covered with a mask (right)	21
Figure 6 – Geometrical evaluation using the GD pattern	22
Figure 7 – Visual evaluation of viewing angle response	23
Figure 8 – Example of the measured LUMINANCE in relation to the standard LUMINANCE response function according to GREYSCALE STANDARD DISPLAY FUNCTION (GSDF).....	26
Figure 9 – An example of the CONTRAST response computed from 18 grey levels as related to the expected CONTRAST response associated with the DICOM 3.14 [2] standard LUMINANCE response with a given tolerance limit (e.g. 15 %) [10]	26
Figure B.1 – Method A, telescopic method.....	48
Figure B.2 – Method B, near-range LUMINANCE meter in combination with an ILLUMINANCE meter	49
Figure B.3 – Method C, frontal integrated LUMINANCE meter in combination with ILLUMINANCE meter	49
Figure B.4 – Method D, back integrated LUMINANCE meter in combination with ILLUMINANCE meter	50
Figure C.1 – Example of TG-18 QC pattern for a matrix size of 1536 × 2048	59
Table 1 – Overview to the definitions of physical parameters	11
Table 2 – TEST PATTERNS used for display testing	14
Table 3 – List of the evaluation methods that can be used for testing medical IMAGE DISPLAY SYSTEMS	16
Table A.1 – Acceptance test sample report of a diagnostic display.....	30
Table A.2 – Constancy test sample report of a diagnostic display	35
Table A.3 – Acceptance test sample report of a monochrome reviewing display	38
Table A.4 – Constancy test sample report of a monochrome reviewing display	40
Table A.5 – Acceptance test sample report of a colour reviewing display	42
Table A.6 – Constancy test sample report of a colour reviewing display	45
Table C.1 – Description of multi-purpose TEST PATTERNS.....	52
Table C.2 – TG18-QC pattern: LUMINANCE levels with 8-bit and [12-bit] pixel values and CX ratings.....	55
Table C.3 – The blurring characteristics of the CX reference set utilized in TG18-QC TEST PATTERNS [16]	56
Table C.4 – Evaluation criteria for the examples of the CLINICAL REFERENCE IMAGES	57
Table C.5 – Example description of TG-18 QC pattern for a matrix size of 1536 × 2048	58

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –
MEDICAL IMAGE DISPLAY SYSTEMS –****Part 1: Evaluation methods**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

DISCLAIMER

This Consolidated version is not an official IEC Standard and has been prepared for user convenience. Only the current versions of the standard and its amendment(s) are to be considered the official documents.

This Consolidated version of IEC 62563-1 bears the edition number 1.1. It consists of the first edition (2009-12) [documents 62B/743/CDV and 62B/768/RVC] and its amendment 1 (2016-03) [documents 62B/983/CDV and 62B/995/RVC]. The technical content is identical to the base edition and its amendment.

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

International Standard IEC 62563-1 has been prepared by subcommittee 62B: Diagnostic imaging equipment of technical committee 62: Electrical equipment in medical practice.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THIS INTERNATIONAL STANDARD, OR AS NOTED: SMALL CAPITALS.

A list of all parts of the IEC 62563 series, published under the general title *Medical electrical equipment – Medical image display systems*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This International Standard provides evaluation methods for testing IMAGE DISPLAY SYSTEMS used in MEDICAL ELECTRICAL EQUIPMENT and medical electrical systems for diagnostic imaging.

On site or after installation, two types of testing can be carried out. An acceptance test is carried out after a new IMAGE DISPLAY SYSTEM has been installed, or major modifications have been made to the existing IMAGE DISPLAY SYSTEM. Since an IMAGE DISPLAY SYSTEM may degrade over time, the constancy test is carried out by the user in a periodic cycle to verify that the performance is maintained for the intended use.

The standard describes various evaluation methods without dictating what particular tests shall be used for acceptance and/or constancy tests.

Rather, it is the intention of this standard to be a reference for other standards and guidelines specific to each modality or to be defined by national authorities who will refer to the evaluation methods of this standard and mention limiting values and frequencies for acceptance and constancy tests. Annex A shows sample reports of such a reference.

To maintain the homogeneity in the IEC standards for MEDICAL ELECTRICAL EQUIPMENT, IEC 61223-2-5, *Evaluation and routine testing in medical imaging departments – Part 2-5: Constancy tests – Image display devices* should be reviewed.

INTRODUCTION to Amendment 1

This amendment is published to introduce colour measurement.

Since publication of IEC 62563-1:2009, IEC 61223-2-5, *Evaluation and routine testing in medical imaging departments Part 2-5: Constancy tests – Image display devices* has been reviewed and withdrawn.

MEDICAL ELECTRICAL EQUIPMENT – MEDICAL IMAGE DISPLAY SYSTEMS –

Part 1: Evaluation methods

1 Scope

This part of IEC 62563 describes the evaluation methods for testing medical IMAGE DISPLAY SYSTEMS.

The scope of this International Standard is directed to practical tests that can be visually evaluated or measured using basic test equipment. More advanced or more quantitative measurements can be performed on these devices, but these are beyond the scope of this document.

This standard applies to medical IMAGE DISPLAY SYSTEMS, which can display ~~monochrome~~ image information ~~in the form of greyscale values on colour and on~~ greyscale and colour IMAGE DISPLAY SYSTEMS ~~(e.g. CATHODE RAY TUBE (CRT) monitors, FLAT PANEL DISPLAYS, PROJECTION SYSTEM)~~. This standard applies to medical IMAGE DISPLAY SYSTEMS used for diagnostic (interpretation of medical images toward rendering clinical diagnosis) or viewing (viewing medical images for medical purposes other than for providing a medical interpretation) purposes and therefore having specific requirements in terms of image quality. Head mounted IMAGE DISPLAY SYSTEMS and IMAGE DISPLAY SYSTEMS used for confirming positioning and for operation of the system are not covered by this standard. **Handheld IMAGE DISPLAY SYSTEMS might require additional or modified versions of the procedures described in this standard.**

It is not in the scope of this standard to define the requirements of acceptance and constancy tests ~~nor~~ the frequencies of constancy tests.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60788:2004, *Medical electrical equipment – Glossary of defined terms*

ISO 11664-1:2007, *Colorimetry – Part 1: CIE standard colorimetric observers*

CIE S 010/E:2004 *Photometry – The CIE system of physical photometry*

3 Terms, definitions, symbols and abbreviations

3.1 Terms and definitions

For the purpose of this document, the terms and definitions given in IEC 60788:2004 and the following apply.

3.1.1

accuracy

closeness of agreement between a test result and the accepted reference value