

JIS

JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS T 3226-2 : 2015

(MTJAPAN/JSA)

**Needle-based injection systems for
medical use—Part 2: Needles—
Requirements and test methods**

ICS 11.040.25

Reference number : **JIS T 3226-2 : 2015 (E)**

Date of Establishment: 2005-03-25

Date of Revision: 2015-10-01

Date of Public Notice in Official Gazette: 2015-10-01

Investigated by: Japanese Industrial Standards Committee
Standards Board for ISO area
Technical Committee on Medical Equipment

JIS T 3226-2:2015, First English edition published in 2016-05

Translated and published by: Japanese Standards Association
Mita MT Building, 3-13-12, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2016

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

NH/AT

Contents

	Page
Introduction.....	1
1 Scope.....	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	4
4.1 Materials	4
4.2 Dimensions	4
4.3 Determination of flow rate through the needle	4
4.4 Bond between hub and needle tube	4
4.5 Needle points	5
4.6 Freedom from defects	5
4.7 Lubrication	5
4.8 Dislocation of measuring point at patient end	5
4.9 Determination of functional compatibility with needle-based injection systems	5
4.10 Ease of assembly and disassembly	5
4.11 Sterility	5
5 Sampling	6
6 Pre-conditioning of needles	6
6.1 Pre-conditioning in a dry-heat atmosphere	6
6.2 Pre-conditioning in a cold-storage atmosphere	6
6.3 Pre-conditioning in a cyclical atmosphere	6
7 Standard atmosphere and apparatus for tests	7
7.1 General	7
7.2 Standard test atmosphere	7
7.3 Test gauge	7
8 Determination of dislocation of measuring point at patient end	8
9 Bond between hub and needle tube	8
10 Packaging.....	9
11 Test method for validating the compatibility of needles and injector systems	9
11.1 Principle	9
11.2 Apparatus	9
11.3 Sample quantity requirements	10

11.4	Procedure	10
11.5	Acceptance criteria	12
11.6	Test report	13
12	Information supplied by the manufacturer	13
12.1	General	13
12.2	Marking	13
12.3	Instructions for use	15
Annex A (normative)	Determination of flow rate through needle	16
Bibliography	18
Annex JA (informative)	Comparison table between JIS and corresponding International Standard	19

Foreword

This translation has been made based on the original Japanese Industrial Standard revised by the Minister of Health, Labour and Welfare through deliberations at the Japanese Industrial Standards Committee as the result of proposal for revision of Japanese Industrial Standard submitted by Medical Technology Association of Japan (MTJAPAN)/Japanese Standards Association (JSA) with the draft being attached, based on the provision of Article 12 Clause 1 of the Industrial Standardization Law applicable to the case of revision by the provision of Article 14.

Consequently **JIS T 3226-2:2011** is replaced with this Standard.

This **JIS** document is protected by the Copyright Law.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, applications for a patent after opening to the public or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, applications for a patent after opening to the public or utility model rights.

JIS T 3226 series consists of the following 2 parts under the general title “*Needle-based injection systems for medical use*”:

Part 1: Needle-based injection systems—Requirements and test methods

Part 2: Needles—Requirements and test methods

Needle-based injection systems for medical use—Part 2: Needles— Requirements and test methods

Introduction

This Japanese Industrial Standard has been prepared based on the second edition of **ISO 11608-2** published in 2012 with some modifications of the technical contents so as to reflect the actualities in Japan.

The portions given dotted underlines are the matters in which the contents of the corresponding International Standard have been modified. A list of modifications with the explanations is given in Annex JA.

1 Scope

This Standard specifies requirements and test methods for single-use, double-ended, sterile needles for needle-based injection systems (NISs) that fulfil the specifications of **JIS T 3226-1** (hereafter referred to as “needles”). It is not applicable to:

- needles for dental use;
- pre-filled syringe needles;
- needles pre-assembled by the manufacturer;
- needles not requiring assembly or attachment to the NIS.

NOTE 1 **JIS T 3226-2:2011** may be applied until 30 September, 2018.

NOTE 2 The International Standard corresponding to this Standard and the symbol of degree of correspondence are as follows.

ISO 11608-2:2012 *Needle-based injection systems for medical use—Requirements and test methods—Part 2: Needles* (MOD)

In addition, symbols which denote the degree of correspondence in the contents between the relevant International Standard and **JIS** are IDT (identical), MOD (modified), and NEQ (not equivalent) according to **ISO/IEC Guide 21-1**.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this Standard. For standards with the year indication, only the editions of the indicated year shall be applied and the revisions (including amendments) made thereafter shall not be applied. For those without the indication of the year, the most recent edition (including amendments) shall be applied.

JIS C 60068-2-30:2011 *Environmental testing—Part 2-30: Tests—Test Db: Damp heat, cyclic (12 h + 12 h cycle)*

NOTE : Corresponding International Standard: IEC 60068-2-30:2005 *Environmental testing—Part 2-30: Tests—Test Db: Damp heat, cycle (12 h + 12 h cycle)* (IDT)