

Institute of Environmental Sciences and Technology

IEST-RP-DTE011.2

Design, Test, and Evaluation Division
Recommended Practice 011.2

**Mechanical Shock and
Vibration Accelerometer
Selection**

Arlington Place One
2340 S. Arlington Heights Road, Suite 620
Arlington Heights, IL 60005-4510
Phone: (847) 981-0100 • Fax: (847) 981-4130
E-mail: information@iest.org • Web: www.iest.org

This Recommended Practice was prepared by and is under the jurisdiction of Working Group 011 of the IEST Design, Test, and Evaluation Division (WG-DTE011). The following WG voting members contributed to the development of this edition of this Recommended Practice.

Jon Wilson, DTE-011 Chair, Jon S. Wilson Consulting, LLC

David Banaszak, Air Force Research Laboratory

Vesta Bateman, Mechanical Shock Consulting

Anthony Chu, TE Connectivity

Randy Patrick, Army Yuma Test Center

Stanley Poynor, Lockheed Martin Missiles & Fire Control

David Smallwood, Consultant

Patrick Walter, TCU Department of Engineering

The Working Group also wishes to acknowledge the contributions of members who prefer to remain anonymous.

Copyright © 2016 by the Institute of Environmental Sciences and Technology

First printing, August 2016

ISBN: 978-1-937280-33-8 (Electronic format)

ISBN: 978-1-937280-32-1 (Print format)

PROPOSAL FOR IMPROVEMENT: The Working Groups of the Institute of Environmental Sciences and Technology are continually working on improvements to their Recommended Practices and Reference Documents. Suggestions from users of these documents are welcome. If you have a suggestion regarding this document, please use the online Proposal for Improvement form found on the IEST website at www.iest.org.

Institute of Environmental Sciences and Technology
Arlington Place One
2340 S. Arlington Heights Road, Suite 620
Arlington Heights, IL 60005-4510
Phone: (847) 981-0100 • Fax: (847) 981-4130
E-mail: information@iest.org • Web: www.iest.org

NOTICE AND DISCLAIMER: This Recommended Practice is published by the Institute of Environmental Sciences and Technology (IEST) to advance the technical and engineering sciences. Use of this document is entirely voluntary, and determination of its applicability and suitability for any particular use is solely the responsibility of the user. Use of this Recommended Practice does not imply any warranty or endorsement by IEST.

The information in this publication was considered technically sound by the consensus of persons engaged in the development and approval of the document at the time it was developed. Consensus does not necessarily mean that there is unanimous agreement among all persons participating in the development of this document.

IEST standards, recommended practices, and guideline publications, of which the document contained herein is one, are developed through a voluntary consensus standards development process. This process brings together volunteers and seeks out the views of persons who have an interest in the topic covered by this publication. While IEST administers the process to promote fairness in the development of consensus, the organization does not write the document and does not independently test, evaluate, or verify the accuracy or completeness of any information or the soundness of any judgments contained in IEST standards, recommended practices, and guideline publications.

IEST disclaims liability for any personal injury, property, or other damages of any nature whatsoever, whether special, indirect, consequential, or compensatory, directly or indirectly resulting from the publication, use of, application, or reliance on this document. IEST disclaims and makes no guaranty or warranty, express or implied, as to the accuracy or completeness of any information published herein, and disclaims and makes no warranty that the information in this document will fulfill any of the particular purposes or needs of users of the document. IEST does not guarantee the performance of any individual manufacturer or seller's products or services by virtue of this standard or guide.

In publishing and making this document available, IEST is not undertaking to render professional or other services for or on behalf of any person or entity, nor is IEST undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances. Information and other standards on the topic covered by this publication may be available from other sources, which the user may wish to consult for additional views or information not covered by this publication.

IEST has no power, nor does it undertake to police or enforce compliance with the contents of this document. IEST does not certify, test, or inspect products, designs, or installations for safety or health purposes. Any certification or other statement of compliance with any health or safety-related information in this document shall not be attributable to IEST and is solely the responsibility of the certifier or maker of the statement.

Mechanical Shock and Vibration Accelerometer Selection

IEST-RP-DTE011.2

CONTENTS

SECTION

1	SCOPE AND LIMITATION	5
2	REFERENCES	5
3	TERMS AND DEFINITIONS	5
4	BACKGROUND AND PURPOSE	10
5	ACCELEROMETER TYPES FOR SHOCK AND VIBRATION	11
6	DYNAMIC CHARACTERISTICS	24
7	ENVIRONMENTAL CHARACTERISTICS	28
8	ACCELEROMETER INSTALLATION	31
9	ACCELEROMETER RECALIBRATION	34
10	ACCELEROMETER SELECTION BY APPLICATION [21]	35

FIGURES

1	COMPRESSION AND SHEAR CRYSTAL STRAINS	11
2	CLASSIFICATIONS OF CRYSTAL TYPES	12
3	EQUIVALENT CIRCUIT OF A PE SENSOR AND A CHARGE AMPLIFIER	13
4	COMMON PE ACCELEROMETER DESIGNS: COMPRESSION TYPE (LEFT); SHEAR TYPE (RIGHT)	15
5	SCHEMATICS OF AN IEPE ACCELEROMETER	15
6	FREQUENCY RESPONSE DROOP	15
7	WHEATSTONE BRIDGE	19
8	MONOLITHIC SILICON SENSING ELEMENT OF A PR ACCELEROMETER	20
9	BLOCK DIAGRAM OF A VC ACCELEROMETER	21
10	EXPLODED VIEW OF A MICROMACHINED VC ACCELEROMETER	22
11	PHASE RESPONSE OF A SINGLE-DOF SYSTEM VS. NORMALIZED FREQUENCY	26
12	TRANSVERSE SENSITIVITY	27
13	ADHESIVE MOUNTING	33
14	MAGNETIC MOUNTING	33
15	ANGULAR ACCELEROMETER USING A LINEAR PAIR	41

TABLE

1	COMPARISON OF PE, IEPE, PR, AND VC ACCELEROMETERS	23
---	---	----

APPENDIX

A	REFERENCES AND BIBLIOGRAPHY	42
---	-----------------------------------	----

Mechanical Shock and Vibration Accelerometer Selection

IEST-RP-DTE011.2

1 SCOPE AND LIMITATION

The purpose of this Recommended Practice (RP) is to provide guidelines for selecting accelerometers to measure shock and vibration in laboratory and field testing environments. Many special applications are not covered (e.g., pyroshock [1,2], consumer products) because of their unique nature and the rapid advancements taking place in their disciplines. Even in applications not specifically addressed, however, these recommendations may be helpful.

There are basically two classes of motion sensors: fixed-reference and mass-spring (relative motion). Non-contact transducers, such as laser interferometric displacement and laser Doppler velocity transducers, are fixed-reference designs. Although they offer some unique properties, these instruments are used to measure shock and vibration only in applications where a fixed reference is available, and where their cost, size, and physical space and geometry requirements are acceptable. Similarly, video and high-speed photographic displacement measurement techniques are becoming more sophisticated, thereby increasingly allowing their application to the motion analysis of structures. These fixed-reference techniques, which have different constraints, are discussed elsewhere [3,4,5]. This RP concentrates on the more common mass-spring type accelerometers, with the sensing element(s) represented by the spring.

The following recommendations apply to dynamic measurements with frequencies ranging from DC (0 Hz) to more than 20 kHz. Only measurements of linear (translational) motion are considered; measurement of angular or rotational motion is addressed as an application at the end of the document.

2 REFERENCES

It is recommended that the user reference manufacturers' specification and application data in the selection and use of equipment. Please see Appendix A: Bibliography for informative resources.

3 TERMS AND DEFINITIONS

accelerometer

A transducer whose instantaneous output is proportional to the instantaneous acceleration input.

acoustic sensitivity

The rated output produced by a non-acoustic transducer in the presence of a specified acoustic field.

back-to-back comparison method

A method of performing a sensitivity/frequency response calibration of an accelerometer by mounting the unit under test to the sensitive surface of a reference standard and comparing the outputs of the two devices.

base strain sensitivity

The rated output produced by an accelerometer in the presence of a specified amount of strain input induced by the bending motion at the mounting interface.