

QMS24

Using Proficiency Testing and Alternative Assessment to Improve Medical Laboratory Quality

This guideline describes an approach for a complete proficiency testing (PT) process and provides assistance to laboratories in using PT as a quality improvement tool.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeals Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeals, documented in the CLSI Standards Development Policies and Processes, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute
950 West Valley Road, Suite 2500
Wayne, PA 19087 USA
P: +1.610.688.0100
F: +1.610.688.0700
www.clsi.org
standard@clsi.org

Using Proficiency Testing and Alternative Assessment to Improve Medical Laboratory Quality

Teresa Darcy, MD, MMM
Gary L. Horowitz, MD
Deirdre Astin, MS, MT(ASCP)
Lucia M. Berte, MA, MT(ASCP)SBB,
DLM, CQA(ASQ)CMQ/OE
Joanne Brisbois
Shilo Caulfield, BS, CQIA (ASQ)
Kathryn Connolly, CQA(ASQ), MT(ASCP)
Gulderen Yanikkaya Demirel, MD, PhD

Bruce Jones, MD
Bob Kaplanis, PBT, MT(ASCP)
Christopher Lyle, MS, MBA
Penny Petinos, BS, MBus
Alan T. Remaley, MD, PhD
Heather Stang, MS
Daniel W. Tholen, MS
Nico Vandepoele
Min Xu, MD, PhD

Abstract

Clinical and Laboratory Standards Institute guideline QMS24—*Using Proficiency Testing and Alternative Assessment to Improve Medical Laboratory Quality* provides laboratories with a detailed description of important activities in the proficiency testing (PT) process and includes suggestions for how to improve this process from a quality management perspective. It includes a suggested classification of unacceptable PT results and specific examples of investigations of unacceptable results.

Clinical and Laboratory Standards Institute (CLSI). *Using Proficiency Testing and Alternative Assessment to Improve Medical Laboratory Quality*. CLSI guideline QMS24 (ISBN 1-56238-944-0 [Print]; ISBN 1-56238-945-9 [Electronic]). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2016.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org.

If you or your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at:

P: +1.610.688.0100 **F:** +1.610.688.0700 **E:** customerservice@clsi.org **W:** www.clsi.org

Copyright ©2016 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedures manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. *Using Proficiency Testing and Alternative Assessment to Improve Medical Laboratory Quality*. 3rd ed. CLSI guideline QMS24. Wayne, PA: Clinical and Laboratory Standards Institute; 2016.

Previous Editions:

GP27: April 1997, August 1999, February 2007
GP29: July 2001, December 2002, August 2008

ISBN 1-56238-944-0 (Print)

ISBN 1-56238-945-9 (Electronic)

ISSN 1558-6502 (Print)

ISSN 2162-2914 (Electronic)

Volume 36, Number 11

Committee Membership

Consensus Council

Carl D. Mottram, RRT, RPFT, FAARC
Chairholder
Mayo Clinic
USA

J. Rex Astles, PhD, FACB, DABCC
 Centers for Disease Control and
 Prevention
 USA

Lucia M. Berte, MA, MT(ASCP)SBB,
 DLM, CQA(ASQ)CMQ/OE
 Laboratories Made Better!
 USA

Karen W. Dyer, MT(ASCP), DLM
 Centers for Medicare & Medicaid
 Services
 USA

Dennis J. Ernst, MT(ASCP), NCPT(NCCT)
 Center for Phlebotomy Education
 USA

Thomas R. Fritsche, MD, PhD, FCAP,
 FIDSA
 Marshfield Clinic
 USA

Mary Lou Gantzer, PhD, FACB
 BioCore Diagnostics
 USA

Loralie J. Langman, PhD
 Mayo Clinic
 USA

Joseph Passarelli
 Roche Diagnostics Corporation
 USA

James F. Pierson-Perry
 Siemens Healthcare Diagnostics Inc.
 USA

Andrew Quintenz
 Bio-Rad Laboratories, Inc.
 USA

Robert Rej, PhD
 New York State Department of
 Health – Wadsworth Center
 USA

Zivana Tezak, PhD
 FDA Center for Devices and
 Radiological Health
 USA

Document Development Committee on Using Proficiency Testing to Improve the Clinical Lab

Teresa Darcy, MD, MMM
Chairholder
University of Wisconsin Health
USA

Gary L. Horowitz, MD
Vice-Chairholder
Beth Israel Deaconess Medical Center
USA

Penny Petinos, BS, MBus
Committee Secretary
RCPA Quality Assurance Programs Pty
Limited
Australia

Bruce Jones, MD
 College of American Pathologists
 USA

Bob Kaplanis, PBT, MT(ASCP)
 Laboratory Sciences of Arizona
 USA

Heather Stang, MS
 Centers for Disease Control and
 Prevention
 USA

Daniel W. Tholen, MS
 American Association for Laboratory
 Accreditation
 USA

Nico Vandepoele
 Bio-Rad Laboratories, Inc.
 USA

Staff

Clinical and Laboratory Standards
 Institute
 USA

Marcy L. Hackenbrack, MCM, M(ASCP)
Project Manager

Megan L. Tertel, MA, ELS
Editorial Manager

Joanne P. Christopher, MA, ELS
Editor

Laura Martin
Editor

Michael A. Russell, MA
Editor

Acknowledgment

CLSI, the Consensus Council, and the Document Development Committee on Using Proficiency Testing to Improve the Clinical Lab gratefully acknowledge the following volunteers for their important contributions to the development of this guideline:

Deirdre Astin, MS, MT(ASCP)
New York State Department of Health
USA

Gulderen Yanikkaya Demirel, MD, PhD
Yeditepe University School of Medicine
Turkey

Lucia M. Berte, MA, MT(ASCP)SBB, DLM,
CQA(ASQ)CMQ/OE
Laboratories Made Better!
USA

Christopher Lyle, MS, MBA
Pel-Freez Arkansas, LLC
USA

Joanne Brisbois
Alberta Health Services
Canada

Alan T. Remaley, MD, PhD
National Institutes of Health
USA

Shilo Caulfield, BS, CQIA (ASQ)
Viracor-IBT Reference Laboratory
USA

Min Xu, MD, PhD
Seattle Children's Hospital
USA

Kathryn Connolly, CQA(ASQ), MT(ASCP)
COLA
USA

Contents

Abstract	i
Committee Membership	iii
Foreword	vii
Chapter 1: Introduction	1
1.1 Scope	2
1.2 Background	2
1.3 Standard Precautions	3
1.4 Terminology	3
Chapter 2: The Proficiency Testing Process	13
2.1 Proficiency Testing Needs Assessment	18
2.2 Development of the Laboratory's Overall Proficiency Testing Plan	18
Chapter 3: Options for Proficiency Testing	19
3.1 Provider-Based Proficiency Testing Process	20
3.2 Alternative Assessment Procedures	28
Chapter 4: Review of Proficiency Testing Reports	33
4.1 Timely Review of Reports	34
4.2 Acceptable Performance (Including Opportunities for Improvement)	34
4.3 Unacceptable Performance	35
4.4 Unscored Results	42
4.5 Opportunities for Improvement and Corrective Action	44
Chapter 5: Results Recorded, Communicated, and Retained	47
5.1 Communication	48
5.2 Records Retention	48
Chapter 6: Ongoing Performance Review and Assessment of the Proficiency Testing Process	51
Chapter 7: Other Uses for Proficiency Testing Samples and Reports	53
7.1 Staff Training and Competence Assessment	54
7.2 Information on Instruments or Methods	55
7.3 Information on Current Practice	56
7.4 Method Evaluation	56

Contents (Continued)

Chapter 8: Special Circumstances for Proficiency Testing in Specific Laboratory Disciplines	59
8.1 Transfusion Medicine	60
8.2 Microbiology	60
8.3 Laboratory-Developed Tests	61
8.4 Molecular Methods	62
8.5 Multiple Instruments and/or Methods	63
8.6 Situations Where Proficiency Testing Is Not Required	65
8.7 Anatomical Pathology	66
8.8 Using Proficiency Testing to Improve Quality in Gynecological Cytology	67
Chapter 9: Conclusion	73
Chapter 10: Supplemental Information	75
References	76
Appendix A. Examples of Worksheets for Tracking Proficiency Testing	84
Appendix B. Examples of Attestation Statements	86
Appendix C. Proficiency Testing Provider Performance Statistics for Quantitative Results	88
Appendix D. Statistical Evaluation of Qualitative Split-Sample Results	90
Appendix E. Procedure to Determine Allowable Differences Between Laboratories X and Y in a Split-Sample Program	91
Appendix F. Examples of Proficiency Testing Results	96
Appendix G. Proficiency Testing Troubleshooting Guide	101
Appendix H. Examples of Proficiency Testing Result Review, Evaluation, and Investigation Forms	105
Appendix I. Molecular Proficiency Testing Report Examples	112
Appendix J. Sample Form for Documenting Unsatisfactory Gynecological Cytology Proficiency Testing Performance	114
The Quality Management System Approach	116
Related CLSI Reference Materials	118

Foreword

Proficiency testing (PT) is a valuable tool in the quality improvement process. PT provides one measure of objective evidence of laboratory competence to customers and regulatory and accreditation organizations. It serves as a unique source of information not obtainable by other methods. QMS24 provides guidance to laboratories on designing a PT process and using PT results, whether acceptable or unacceptable, to improve the quality of laboratory testing. PT cannot be used as the sole means for evaluating the quality of a laboratory, as PT is only one component of laboratory quality management. Current accreditation requirements include integration of PT into the laboratory's quality improvement program, and this guideline describes how that can be accomplished.

Overview of Changes

This guideline replaces the second edition of GP27, published in 2007, and has been recoded as QMS24. Several changes were made in this edition, including:

- ▶ The terminology and definitions were updated and clarified.
- ▶ The scope of the guideline was expanded to include information published in CLSI document GP29, and to eliminate redundancy with that document.
- ▶ The entire guideline was reorganized and updated to be consistent with CLSI's quality system essentials, with a focus on using a process workflow for the PT process.
- ▶ A process flow chart was added that outlines development, implementation, and monitoring of the PT process.
- ▶ Additional information on opportunities for improvement for laboratories in longitudinal review of successful PT events was included.
- ▶ Additional information to assist laboratories in using PT to assess and improve laboratory quality was included.
- ▶ Chapters were added to provide an in-depth discussion of PT in specialized areas of the laboratory, such as molecular and gynecological cytology.

KEY WORDS

Alternative assessment procedure

Corrective action

External quality assessment

Proficiency testing

Quality assurance

Quality improvement

This page is intentionally left blank.

Chapter 1

Introduction

This chapter includes:

- ▶ Guideline’s scope and applicable exclusions
- ▶ Background information pertinent to the guideline’s content
- ▶ Standard precautions information
- ▶ “Note on Terminology” that highlights particular use and/or variation in use of terms and/or definitions
- ▶ Terms and definitions used in the guideline
- ▶ Abbreviations and acronyms used in the guideline



QUALITY

Using Proficiency Testing and Alternative Assessment to Improve Medical Laboratory Quality

1 Introduction

1.1 Scope

IMPORTANT NOTE:

Regulatory and accreditation organizations may have additional requirements extending beyond the guidance in QMS24.

REMINDER:

See CLSI document QMS11⁷ for information on how to conduct a root cause analysis.

NOTE:

PT serves as an external verification of a laboratory's results, and also as a valuable self-monitoring tool.

The purpose of this guideline is to help medical laboratories use proficiency testing (PT) as a quality improvement tool.¹⁻⁶ This guideline presents a systematic approach for designing the PT process as a component of the laboratory QMS.

QMS24 is intended for clinical laboratory managers and analysts in both the public and private sectors, and is applicable to any setting in which clinical laboratory testing is performed, from bedside testing to large multispecialty laboratories. This guideline applies to both qualitative and quantitative laboratory testing, including detection and quantification of blood and fluid measurands and blood and tissue typing. Some discussions apply only to examinations with quantitative results, whereas other discussions apply to examinations with qualitative results.

The processes described in this guideline can help laboratories design a PT process, monitor PT results, and investigate and respond to unacceptable PT results. Part of this response may include preparation of information for submission to regulatory or accreditation organizations. Laboratories are cautioned, however, that regulatory and accreditation organizations may have additional requirements not supported by the guidance in QMS24.

QMS24 also provides guidance for how to use PT as a tool to prevent problems through analysis of acceptable results, education of laboratory personnel, and monitoring of internal processes.

This guideline does not recommend specific corrective actions for specific root causes (see CLSI document QMS11⁷).

1.2 Background

PT evaluates a laboratory's performance on various types of testing and examinations in comparison to peer group performance or a reference standard or method. Alternative assessment procedures (AAPs) may evaluate testing and examination performance against a reference laboratory or against clinical information. PT serves as an external verification of a laboratory's results, and also as a valuable self-monitoring tool. PT directly benefits the laboratory and, indirectly, its customers and regulatory and accreditation organizations.

The use of PT to improve the quality of laboratory performance is not limited to the investigation of unacceptable results. Monitoring