

FINAL VERSION

VERSION FINALE



**Electroacoustics – Hearing aids –
Part 4: Induction-loop systems for hearing aid purposes – System performance
requirements**

**Électroacoustique – Appareils de correction auditive –
Partie 4: Systèmes de boucles d'induction utilisées à des fins de correction
auditive – Exigences de performances système**

CONTENTS

FOREWORD.....	6
INTRODUCTION.....	8
1 Scope.....	9
2 Normative references	9
3 Terms and definitions	9
4 General	10
4.1 Procedure for setting up and commissioning an audio-frequency induction loop system	10
4.2 Suitability of the site for the installation of an audio-frequency induction-loop system.....	11
4.3 Relation of the magnetic field strength level at the telecoil to the sound pressure level at the microphone.....	11
5 Using components of a sound system in an induction-loop system	11
5.1 General.....	11
5.2 Microphones	11
5.3 Mixer	11
5.4 Power amplifier.....	11
6 Meters and test signals.....	12
6.1 Meters	12
6.1.1 Meters in general.....	12
6.1.2 Requirements common to both types	12
6.1.3 True-r.m.s. meter.....	12
6.1.4 Peak programme meter (PPM).....	12
6.2 Test signals in general.....	12
6.3 Speech signals	13
6.3.1 Live speech signals	13
6.3.2 Recorded speech material	13
6.3.3 Simulated speech material.....	13
6.4 Pink noise signal.....	14
6.5 Sinusoidal signal.....	14
6.6 Combi signal.....	14
7 Magnetic background noise level of the installation site	15
7.1 Method of measurement	15
7.2 Recommended maximum magnetic noise levels.....	15
8 Characteristics to be specified, methods of measurement and requirements.....	16
8.1 General.....	16
8.2 Magnetic field strength.....	17
8.2.1 Characteristic to be specified.....	17
8.2.2 Method of measurement with a simulated speech signal.....	17
8.2.3 Method of measurement with pink noise	17
8.2.4 Method of measurement with a sinusoidal signal	17
8.2.5 Method of measurement with a combi signal.....	17
8.2.6 Method of measurement – Other.....	17
8.2.7 Requirements	17
8.3 Frequency response of the magnetic field.....	18
8.3.1 Characteristic to be specified.....	18

8.3.2	Method of measurement with a simulated speech signal.....	18
8.3.3	Method of measurement with pink noise	18
8.3.4	Method of measurement with a sinusoidal signal	19
8.3.5	Method of measurement with combi signal.....	19
8.3.6	Method of measurement – Other.....	19
8.3.7	Requirements	19
8.4	Useful magnetic field volume	19
8.4.1	Characteristic to be specified.....	19
8.4.2	Methods of measurement	20
8.4.3	Requirements	20
9	Small-volume systems	20
9.1	Definition of measurement points	20
9.2	Disabled refuge and similar call-points.....	20
9.3	Counter systems	22
9.4	Useful magnetic field volume method.....	24
9.5	Requirements	24
10	Setting up (commissioning) the system.....	24
10.1	Procedure	24
10.2	Requirements	24
10.3	Amplifier overload at the upper bound of maximum power bandwidth.....	25
10.3.1	Explanation	25
10.3.2	Methods of test.....	25
10.3.3	Requirements	26
10.4	Magnetic noise level due to the system	26
10.4.1	Explanation of term.....	26
10.4.2	Method of measurement with a speech signal.....	26
10.4.3	Method of measurement with pink noise	26
10.4.4	Method of measurement with a sinusoidal signal	26
10.4.5	Method of measurement with a combi signal.....	26
10.4.6	Method of measurement – Other (no input signal).....	27
10.4.7	Requirements	27
Annex A (informative)	Systems for small useful magnetic field volumes	28
A.1	Overview.....	28
A.2	Body-worn audio systems	28
A.3	Small volume, defined seating, mainly in households.....	28
A.4	Specific locations such as help and information points, ticket and bank counters, etc.....	28
Annex B (informative)	Measuring equipment	31
B.1	Overview.....	31
B.2	Signal sources	31
B.2.1	Real speech	31
B.2.2	Simulated speech	31
B.2.3	Pink noise.....	31
B.2.4	Sine wave.....	31
B.3	Magnetic field strength level meter.....	32
B.3.1	General recommendations	32
B.3.2	Peak-programme meter (PPM) type	32
B.3.3	True r.m.s. meter type	32
B.4	Field strength level meter calibrator	33

B.5	Spectrum analyzer	33
Annex C (informative)	Provision of information.....	34
C.1	General.....	34
C.2	Information to be provided to the hearing aid user	34
C.3	Information to be provided to system installers and by them to users	35
C.4	Information to be provided by the manufacturer of the amplifying equipment.....	35
Annex D (informative)	Measuring speech signals	36
Annex E (informative)	Basic theory and practice of audio-frequency induction-loop systems	37
E.1	Properties of the loop and its magnetic field.....	37
E.2	Directional response of the telecoil of a hearing aid	38
E.3	Supplying the loop current	43
E.4	Signal sources and cables	44
E.4.1	Microphones	44
E.4.2	Other signal sources.....	45
E.4.3	Cables	45
E.5	Care of the system.....	45
E.6	Magnetic units	45
Annex F (informative)	Effects of metal in the building structure on the magnetic field.....	46
Annex G (informative)	Calibration of field-strength meters	48
Annex H (informative)	Effect of the aspect ratio of the loop on the magnetic field strength	50
H.1	Overview.....	50
H.2	Effect of aspect ratio on field patterns	50
Annex I (informative)	Overspill of magnetic field from an induction-loop system	53
I.1	General.....	53
I.2	Examples of overspill issues	53
I.3	Addressing overspill issues	53
Bibliography	55
Figure 1	– Flow chart for the operations in this standard	10
Figure 2	– Measurement points for disabled refuge and similar call-points	21
Figure 3	– Measurement points for a counter system	23
Figure A.1	– Field pattern of a vertical loop.....	29
Figure A.2	– Contour plot of field strength of vertical loop	30
Figure C.1	– Graphical symbol: inductive coupling	34
Figure E.1	– Perspective view of a loop, showing the magnetic field vector paths	38
Figure E.2	– Strengths of the components of the magnetic field due to current in a horizontal rectangular loop at points in a plane above or below the loop plane.....	39
Figure E.3	– Field patterns of the vertical component of the magnetic field of a horizontal loop	40
Figure E.4	– Field patterns of the vertical component of the magnetic field of a vertical loop 0,75 m square	41
Figure E.5	– Perspective view of the variation of the vertical field strength level at an optimum height above a horizontal rectangular loop.....	42
Figure E.6	– Directional response of the magnetic pick-up coil (telecoil) of a hearing aid	43
Figure F.1	– Magnetic field pattern of a 10 m by 14 m loop, 1,2 m above its plane	46

Figure F.2 – Magnetic field pattern of a 10 m by 14 m loop, 1,2 m above its plane, showing the effect of metal (iron) in the floor	47
Figure G.1 – Triple Helmholtz coil for calibration of meters	48
Figure H.1 – Variation of the current required to produce a specified magnetic field strength at a specific point with the dimensions and aspect ratio of the loop	50
Figure H.2 – Square and rectangular loops	52
Table 1 – Application of signals	13
Table 2 – Specification of the combi signal	15
Table 3 – Magnetic field strengths typically produced by different test signals, with an amplifier having peak-detecting AGC	18
Table 4 – Programme material and test frequency	26
Table D.1 – Typical variation of the measured RMS voltage of a speech signal with averaging time	36

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**ELECTROACOUSTICS –
HEARING AIDS –****Part 4: Induction-loop systems for hearing aid purposes –
System performance requirements**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

DISCLAIMER

This Consolidated version is not an official IEC Standard and has been prepared for user convenience. Only the current versions of the standard and its amendment(s) are to be considered the official documents.

This Consolidated version of IEC 60118-4 bears the edition number 3.1. It consists of the third edition (2014-12) [documents 29/855/FDIS and 29/861/RVD] and its amendment 1 (2017-11) [documents 29/952/CDV and 29/961/RVC]. The technical content is identical to the base edition and its amendment.

This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.

International Standard IEC 60118-4 has been prepared by IEC technical committee 29: Electroacoustics.

This third edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition: Addition of Annexes G, H and I where more information is provided about practical considerations and methods of measurement.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 60118 series, published under the general title *Electroacoustics – Hearing aids*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

Audio-frequency induction-loop systems are widely used to provide a means for hearing aid users, whose hearing aids are fitted with induction pick-up coils, generally known as 'telecoils', to minimise the problems of listening when at a distance from a source of sound, shielded from the person speaking by a protective window, and/or in a background noise. Background noise and distance are two of the main causes of hearing aid users being unable to hear satisfactorily in other than face-to-face quiet conditions. Induction-loop systems have been widely installed in churches, theatres and cinemas, for the benefit of hearing-impaired people. The use of induction-loop systems has been extended to many transient communication situations such as ticket offices, bank counters, drive-in/drive-through service locations, lifts/elevators etc. The widespread provision of telephone handsets that provide inductive coupling to hearing aids is another significant application, where ITU-T Recommendation P370 [1]¹ applies.

Transmission of an audio-frequency signal via an induction-loop system can often establish an acceptable signal-to-noise ratio in conditions where a purely acoustical transmission would be significantly degraded by reverberation and background noise.

One form of audio frequency induction-loop system comprises a cable installed in the form of a loop usually around the perimeter of a room or area in which a group of hearing impaired persons wish to listen. The cable is connected via an amplifier to a microphone system or other source of audio signal, such as a radio receiver, CD player etc. The amplifier produces an audio-frequency electric current in the induction loop cable, causing a magnetic field to be produced inside the loop. The design and implementation of the induction loop is determined by the construction of the building in which it is installed, particularly by the presence of large amounts of iron, steel or aluminium in the structure. In addition the layout and position of electrical cables and equipment may generate high levels of background audio frequency magnetic fields that may interfere with the reception of the loop signal.

Another form of induction-loop system employs a small loop, intended for communication with a hearing-aid user in its immediate vicinity. Examples are: neck loops, ticket-counter systems, self-contained 'portable' systems and chairs incorporating induction loops. (See Annex A)

The pick-up device for an audio-frequency induction-loop system is usually a personal hearing aid, of a type fitted with a pick-up coil (telecoil); however, special induction loop receivers may be used in certain applications.

¹ Numbers in square brackets refer to the Bibliography.

ELECTROACOUSTICS – HEARING AIDS –

Part 4: Induction-loop systems for hearing aid purposes – System performance requirements

1 Scope

This part of IEC 60118 is applicable to audio-frequency induction-loop systems producing an alternating magnetic field at audio frequencies and intended to provide an input signal for hearing aids operating with an induction pick-up coil (telecoil). Throughout this standard, it is assumed that the hearing aids used with it conform to all relevant parts of IEC 60118.

This standard specifies requirements for the field strength in audio-frequency induction loops for hearing aid purposes, which will give adequate signal-to-noise ratio without overloading the hearing aid. The standard also specifies the minimum frequency response requirements for acceptable intelligibility.

Methods for measuring the magnetic field strength are specified, and information is given on appropriate measuring equipment (see Annex B), information that should be provided to the operator and users of the system (see Annex C), and other important considerations.

This standard does not specify requirements for loop driver amplifiers or associated microphone or audio signal sources, which are dealt with in IEC 62489-1, or for the field strength produced by equipment, such as telephone handsets, within the scope of ITU-T P.370.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60268-3:2013, *Sound system equipment – Part 3: Amplifiers*

IEC 60268-10:1991, *Sound system equipment – Part 10: Peak programme level meters*

IEC 61672-1:2013, *Electroacoustics – Sound level meters – Part 1: Specifications*

IEC 62489-1:2010, *Electroacoustics – Audio-frequency induction-loop systems for assisted hearing – Part 1: Methods of measuring and specifying the performance of system components*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.