

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Nuclear instrumentation – Data format for list mode digital data acquisition used in radiation detection and measurement**

**Instrumentation nucléaire – Format de données pour l'acquisition de données numériques en mode liste utilisées dans la détection et la mesure des rayonnements**



**THIS PUBLICATION IS COPYRIGHT PROTECTED**  
**Copyright © 2018 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

#### **About the IEC**

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

#### **About IEC publications**

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

#### **IEC Catalogue - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)**

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

#### **IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)**

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### **IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)**

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

#### **Electropedia - [www.electropedia.org](http://www.electropedia.org)**

The world's leading online dictionary of electronic and electrical terms containing 21 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### **IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)**

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

#### **IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)**

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

---

#### **A propos de l'IEC**

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

#### **A propos des publications IEC**

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### **Catalogue IEC - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)**

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

#### **Recherche de publications IEC - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)**

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### **IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)**

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

#### **Electropedia - [www.electropedia.org](http://www.electropedia.org)**

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 21 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### **Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)**

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

#### **Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)**

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



---

**Nuclear instrumentation – Data format for list mode digital data acquisition used in radiation detection and measurement**

**Instrumentation nucléaire – Format de données pour l'acquisition de données numériques en mode liste utilisées dans la détection et la mesure des rayonnements**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

---

ICS 27.120.01

ISBN 978-2-8322-6034-0

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	10
INTRODUCTION.....	12
1 Scope.....	13
2 Normative references .....	13
3 Terms and definitions, abbreviated terms and notations .....	14
3.1 Terms and definitions.....	14
3.2 Abbreviated terms.....	18
3.3 Notations .....	18
3.3.1 Graphical notation .....	18
3.3.2 ASN.1 notation .....	18
3.3.3 Decimal separator .....	19
3.3.4 Hexadecimal and binary numbers .....	19
4 General requirements .....	19
4.1 How to use this document.....	19
4.2 The use of Abstract Syntax Notation One (ASN.1) .....	19
4.2.1 General .....	19
4.2.2 Encoding rule .....	19
4.2.3 Extensibility .....	19
4.2.4 Object identifier .....	20
4.2.5 Optional and default elements .....	20
4.3 The concept of list-mode data acquisition .....	20
4.4 Channels and devices in a data file or stream .....	21
4.5 Applicability to data files and streams .....	22
4.6 Relation with IEC 62755 data format files.....	23
4.7 Data security.....	24
5 ASN.1 type definitions .....	24
5.1 Overview.....	24
5.2 ASN.1 types.....	27
5.3 The REAL32 and REAL64 type .....	27
5.3.1 Definition.....	27
5.3.2 Use.....	27
5.4 The REAL32List, REAL64List and INTEGERList types .....	27
5.4.1 Definition .....	27
5.4.2 Use.....	28
5.5 The Numeric type .....	28
5.5.1 Definition.....	28
5.5.2 Use.....	28
5.6 The UTCDateTime .....	28
5.6.1 Definition .....	28
5.6.2 Use.....	28
5.7 The Range type .....	28
5.7.1 Definition .....	28
5.7.2 Use.....	28
5.8 The TimeStamp type.....	29
5.8.1 Definition.....	29
5.8.2 Use.....	29

5.9	The SyncStatus type .....	29
5.9.1	Definition .....	29
5.9.2	Use .....	29
5.10	The FlagKind type .....	29
5.10.1	Definition .....	29
5.10.2	Use .....	29
5.11	The AccuracyKind type .....	30
5.11.1	Definition .....	30
5.11.2	Use .....	30
5.12	The Listmodedata type .....	30
5.12.1	Definition .....	30
5.12.2	Use .....	30
5.13	The Header type .....	31
5.13.1	Definition .....	31
5.13.2	Use .....	31
5.13.3	The standardID element .....	32
5.13.4	The listModeDataID element .....	32
5.13.5	The listModeDataPart and listModeDataNParts elements .....	32
5.13.6	The measSetupID element .....	32
5.13.7	The measSetupDescription element .....	32
5.13.8	The iec62755 element .....	32
5.13.9	The radSource element .....	32
5.13.10	The start element .....	33
5.13.11	The startAccuracy element .....	33
5.13.12	The deviceList element .....	33
5.13.13	The channelList element .....	33
5.13.14	The messageList element .....	34
5.14	The IEC62755 type .....	34
5.14.1	Definition .....	34
5.14.2	Use .....	34
5.15	The Device type .....	35
5.15.1	Definition .....	35
5.15.2	Use .....	35
5.15.3	The name, manuf, model and serial elements .....	35
5.16	The Channel type .....	35
5.16.1	Definition .....	35
5.16.2	Use .....	36
5.16.3	The deviceID element .....	36
5.16.4	The kind and physicalChannel elements .....	36
5.16.5	The name element .....	36
5.16.6	The description element .....	36
5.16.7	The parameters element .....	36
5.16.8	The delay element .....	37
5.16.9	The refClock and syncStatus elements .....	37
5.16.10	The adcSamplingRate element .....	37
5.16.11	The adcBitResolution element .....	37
5.16.12	The adcJitterRMS and timeStampJitterRMS elements .....	37

5.16.13	The fineTimeBitResolution element .....	37
5.16.14	The clockFrequency element.....	38
5.16.15	The eventPropertyList element.....	38
5.17	The EventProperty type.....	38
5.17.1	Definition .....	38
5.17.2	Use.....	38
5.18	The EventList type.....	39
5.18.1	Definition .....	39
5.18.2	Use.....	39
5.18.3	The listModeDataID element.....	39
5.18.4	The listModeDataPart element.....	39
5.18.5	The id element .....	39
5.18.6	The eventList element.....	39
5.19	The Event type .....	40
5.19.1	Definition .....	40
5.19.2	Use.....	40
5.20	The EventPulse and EventPulseProperty types .....	40
5.20.1	Definition .....	40
5.20.2	Use.....	40
5.20.3	The channelID element.....	40
5.20.4	The timeStamp element.....	41
5.20.5	The valueList, valueTypeList and valueDescriptionList elements.....	41
5.20.6	The flags, flagDescriptionList and flagKindList elements.....	42
5.20.7	The description element.....	43
5.21	The EventDigitalSignalList and EventDigitalSignalListProperty types .....	43
5.21.1	Definition .....	43
5.21.2	Use.....	43
5.21.3	The channelID element.....	43
5.21.4	The timeStamp element.....	43
5.21.5	The digitalSignalList and digitalSignalPropertyList elements.....	43
5.21.6	The flagDescriptionList and flagKindList elements .....	43
5.21.7	The description element.....	44
5.22	The DigitalSignal and DigitalSignalProperty type.....	44
5.22.1	Definition .....	44
5.22.2	Use.....	44
5.22.3	The samplingRate and defaultSamplingRate element.....	45
5.22.4	The range and defaultRange element.....	45
5.22.5	The delay element .....	45
5.22.6	The samples element.....	45
5.22.7	The flags element .....	46
5.22.8	The description element.....	46
5.22.9	The unit element .....	46
5.22.10	The samplesType element.....	46
5.23	The EventTime and EventTimeProperty types.....	46
5.23.1	Definition .....	46

5.23.2	Use.....	46
5.23.3	The channelID element.....	47
5.23.4	The timeStamp element.....	47
5.23.5	The sinceStart and sincePrevious elements.....	47
5.23.6	The description element.....	47
5.24	The EventRollover and EventRolloverProperty types .....	47
5.24.1	Definition.....	47
5.24.2	Use.....	47
5.24.3	The channelID element.....	48
5.24.4	The nRollovers element.....	48
5.24.5	The description element.....	48
5.24.6	The timeStampCounterBits element .....	48
5.25	The EventRTC and EventRTCProperty types.....	48
5.25.1	Definition .....	48
5.25.2	Use.....	48
5.25.3	The channelID element.....	49
5.25.4	The timeStamp element.....	49
5.25.5	The realtimeclock element.....	49
5.25.6	The syncStatus element.....	49
5.25.7	The description element.....	49
5.26	The EventLogic and EventLogicProperty types .....	49
5.26.1	Definition.....	49
5.26.2	Use.....	50
5.26.3	The channelID element.....	50
5.26.4	The timeStamp element.....	50
5.26.5	The flags, flagDescriptionList and flagKindList elements.....	50
5.26.6	The description element.....	50
5.27	The EventGeo and EventGeoProperty types.....	50
5.27.1	Definition .....	50
5.27.2	Use.....	51
5.27.3	The channelID element.....	51
5.27.4	The timeStamp element.....	51
5.27.5	The position element.....	51
5.27.6	The orientation element.....	51
5.27.7	The speed element .....	51
5.27.8	The description element.....	51
5.28	The Position type.....	52
5.28.1	Definition.....	52
5.28.2	Use.....	52
5.29	The GeographicPoint type.....	52
5.29.1	Definition .....	52
5.29.2	Use.....	52
5.29.3	The latitude element.....	52
5.29.4	The longitude element.....	52
5.29.5	The elevation element.....	53
5.29.6	The elevationOffset element.....	53
5.29.7	The geopointAccuracy element.....	53

5.29.8	The elevationAccuracy element.....	53
5.29.9	The elevationOffsetAccuracy element.....	53
5.29.10	The datum element.....	53
5.30	The RelativeLocation type.....	53
5.30.1	Definition.....	53
5.30.2	Use.....	53
5.30.3	The rellocAzimuth element.....	54
5.30.4	The rellocInclination element.....	54
5.30.5	The distance element.....	54
5.30.6	The origin element.....	54
5.31	The Orientation type.....	55
5.31.1	Definition.....	55
5.31.2	Use.....	55
5.31.3	The azimuth element.....	55
5.31.4	The inclination element.....	55
5.31.5	The roll element.....	56
5.32	The EventMeasurementList and EventMeasurementListProperty types.....	56
5.32.1	Definition.....	56
5.32.2	Use.....	56
5.32.3	The channelID element.....	56
5.32.4	The timeStamp element.....	56
5.32.5	The measurementList element.....	56
5.32.6	The description element.....	57
5.32.7	The measurementPropertyList element.....	57
5.33	The Measurement and MeasurementProperty types.....	57
5.33.1	Definition.....	57
5.33.2	Use.....	57
5.33.3	The value element.....	57
5.33.4	The accuracy and defaultAccuracy elements.....	57
5.33.5	The description element.....	58
5.33.6	The quantity element.....	58
5.33.7	The accuracyKind element.....	58
5.33.8	The unit element.....	58
5.33.9	The valueType element.....	58
5.33.10	The accuracyType element.....	58
5.34	The EventHistogram1DList and EventHistogram1DListProperty types.....	58
5.34.1	Definition.....	58
5.34.2	Use.....	58
5.34.3	The channelID element.....	59
5.34.4	The timeStamp element.....	59
5.34.5	The histogram1DList element.....	59
5.34.6	The description element.....	59
5.34.7	The histogram1DPropertyList element.....	59
5.35	The Histogram1D and Histogram1DProperty types.....	59
5.35.1	Definition.....	59

5.35.2	Use.....	60
5.35.3	The allBins element.....	60
5.35.4	The binList element.....	60
5.35.5	The description element.....	60
5.35.6	The histogramKind element.....	60
5.35.7	The accumulation element.....	61
5.35.8	The binContentType element.....	61
5.35.9	The nBinsX element.....	61
5.35.10	The quantityX and quantityY elements.....	61
5.35.11	The unitX and unitY elements.....	61
5.35.12	The rangeX and rangeY elements.....	61
5.36	The EventHistogram2DList and EventHistogram2DListProperty types.....	61
5.36.1	Definition.....	61
5.36.2	Use.....	61
5.37	The Histogram2D and Histogram2DProperty types.....	62
5.37.1	Definition.....	62
5.37.2	Use.....	62
5.38	The EventMessage and EventMessageProperty types.....	63
5.38.1	Definition.....	63
5.38.2	Use.....	63
5.38.3	The channelID element.....	63
5.38.4	The timeStamp element.....	63
5.38.5	The txtMessage element.....	63
5.38.6	The binMessage element.....	63
5.38.7	The description element.....	64
5.39	The Footer type.....	64
5.39.1	Definition.....	64
5.39.2	Use.....	64
5.39.3	The listModeDataID element.....	64
5.39.4	The listModeDataPart element.....	64
5.39.5	The lastEventListid element.....	64
5.39.6	The stop element.....	64
5.39.7	The totalDeadTimeList and totalLiveTimeList elements.....	64
6	How to use the list-mode data format defined in this document.....	65
6.1	General.....	65
6.2	Using timestamps.....	65
6.3	Dealing with timestamp counter rollover.....	67
6.4	Optimising the size of timestamp and rollover event encodings.....	68
6.5	Specifying high-precision event timing.....	68
6.5.1	General.....	68
6.5.2	Randomisation of *FINE_TIME values to resolve binning issues.....	68
6.6	Specifying events of different types in the same list-mode data.....	69
6.7	Representing data from one device and one channel.....	71
6.8	Encoding/decoding data for writing to/reading from a file or stream.....	73
6.8.1	General requirements.....	73
6.8.2	Writing encoded data to a file.....	73
6.8.3	Writing encoded data to a stream.....	79

6.8.4	Reading encoded data from a file .....	79
6.8.5	Reading encoded data from a stream .....	79
6.9	Representing data from more than one channel .....	80
6.10	Representing data from more than one device .....	82
6.11	Specifying dead time.....	83
6.11.1	Dead time, real time and live time.....	83
6.11.2	Using flags to identify pile-up.....	84
6.11.3	Associating dead time to individual pulses.....	86
6.11.4	Specifying dead time and live/real time not associated to individual pulses.....	87
6.11.5	Specifying different sources of dead time .....	88
6.12	Representing logic events.....	89
6.13	Representing digital signals .....	90
6.14	Representing one-dimensional histograms.....	92
6.15	Representing text or binary data in any format.....	93
6.15.1	General .....	93
6.15.2	Text messages .....	93
6.15.3	Binary or text data in any format .....	93
6.15.4	Sending IEC 62755 data.....	94
Annex A (informative)	Complete ASN.1 syntax for list-mode data .....	95
A.1	ASN.1 syntax .....	95
Annex B (informative)	Introduction to ASN.1 .....	105
B.1	General.....	105
B.2	ASN.1 essentials .....	105
B.3	Advantages of ASN.1 .....	106
B.4	Available tools .....	106
B.5	ASN.1 workflow – an example.....	106
B.5.1	Overview .....	106
B.5.2	ASN.1 specification .....	107
B.5.3	Choice of encoding rule .....	107
B.5.4	Encoding .....	107
B.5.5	Transmission or storage .....	108
B.5.6	Decoding.....	108
Figure 1	– Graphical notations used in this document. ....	18
Figure 2	– Typical data acquisition device controlled by a local computer, which stores list-mode data on a local drive and/or sends it to a remote computer .....	20
Figure 3	– Example of combining channels in one or more data files.....	22
Figure 4	– This document applies to data files and data streams .....	23
Figure 5	– Relations between IEC 63047 and IEC 62755 data files .....	24
Figure 6	– Example with seven values of the <code>Listmodedata</code> type (one value of the <code>Header</code> type, followed by five values of the <code>EventList</code> type and one value of the <code>Footer</code> type).....	26
Figure 7	– Structure of list-mode data files. A rectangle represents the encoding of a value of the <code>Listmodedata</code> type.....	31
Figure 8	– Location of an object in polar coordinates from the reference point .....	54
Figure 9	– Object orientation relative to azimuth, inclination and roll angles.....	55
Figure 10	– Example of timestamp counter rollover.....	67

Figure 11 – Example of plotting 10 000 000 random 10-bit <code>INTEGER</code> values in 1 000 bins, after converting to a real number between 0 and 1 .....	69
Figure 12 – Arrangement of one detector and one data acquisition device, resulting in one channel .....	71
Figure 13 – Example of a data acquisition device providing data from three channels: two radiation detectors and one internal temperature sensor .....	80
Figure 14 – Example where the local computer controls and stores data from two data acquisition devices .....	82
Figure 15 – Dead time associated to individual pulses .....	84
Figure B.1 – The complete encoding of the example <code>value1</code> , using the C-OER encoding rule .....	108
Figure B.2 – Example of transmitting an encoded value .....	109

INTERNATIONAL ELECTROTECHNICAL COMMISSION

NUCLEAR INSTRUMENTATION – DATA FORMAT FOR LIST  
MODE DIGITAL DATA ACQUISITION USED IN RADIATION  
DETECTION AND MEASUREMENT

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 63047 has been prepared by IEC Technical Committee 45: Nuclear instrumentation.

The text of this International Standard is based on the following documents:

CDV	Report on voting
45/842/CDV	45/848/RVC

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Names of data types, elements and definitions in ASN.1: in Courier New

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

**IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

In the context of this document, digital data acquisition devices are high-performance measuring instruments that sample signals from radiation detectors in order to store and process them in a digital manner. They offer advantages over conventional data acquisition systems: they are in general faster and cope better with high count rates, are more compact and less expensive per input channel. They allow easier integration of many radiation detectors in the same system, which results in benefits for nuclear physics. In addition, the real-time processing provided by digital data acquisition systems allows new applications in for example nuclear security and safeguards. They enable off-line processing and re-analysis of acquisition data which may improve the accuracy of radioactivity measurement, and enable computational analysis of individual detector or system response data.

The equipment to which this document applies is commonly known as "digitisers", and may offer signal processing with digital signal processors (DSP) or field-programmable gate arrays (FPGAs). To reduce the amount of data to be stored or transferred, these devices are programmed to extract only the properties of interest, such as the pulse height, pulse area, sub-areas, rise time or pulse shape, and the timestamp of the events associated with the interaction of radiation in a detector. Alternatively, they may also offer the partial or full recording of the digital signal for more advanced processing. This type of time-stamped data is often called "list-mode data" which may be stored for off-line processing and analysis by software.

Typically, digital data acquisition devices are controlled by a computer, which may be embedded in the device, and which runs proprietary or custom-made data acquisition software. Instrument drivers take care of the communication with the data acquisition device. This document specifies the format of the data presented by digital data acquisition devices, primarily used for accurate laboratory measurements, but also for new applications in nuclear safety, security and safeguards. The data may be stored on a local or remote computer, or may be streamed to a remote computer in a network.

This document relies on ASN.1 notation. ASN.1 is an internationally-standardised, vendor-, platform- and language-independent notation to specify data structures at a high level of abstraction. The ASN.1 notation standard is complemented with standardised encoding rules, which are independent of the notation, and again are platform- and language-independent. The encoding rules determine the precise bit-patterns in which values of these data structures are represented, when they are stored in a file or transferred to another computer over a network. Tools are available that facilitate the implementation of this document in software for data acquisition and analysis.

# NUCLEAR INSTRUMENTATION – DATA FORMAT FOR LIST MODE DIGITAL DATA ACQUISITION USED IN RADIATION DETECTION AND MEASUREMENT

## 1 Scope

This international standard specifies the format of binary list-mode data at the output of digital data acquisition devices used for the detection and measurement of radiation. Such data acquisition devices may employ digital signal processors (DSPs) and field-programmable gate arrays (FPGAs) in combination with memory and a communication interface with a computer.

This document is applicable to those data acquisition devices which are able to record and present interaction data of radiation in detectors on an event-per-event basis, with data stored in an output file or streamed to a remote computer. Such list-mode data typically contains timestamp and energy information, but may also contain digital signals or properties like rise time or sub-areas of signals computed by the DSP or FPGA from the signal samples.

It is not the scope of this document to specify the communication protocol between data acquisition hardware and the instrument drivers. Instead, it specifies a standard format for the presentation of data acquired by one or more acquisition devices, to the user or user application.

This document only specifies the binary formatting of data. The formatted data may be sent or stored as specified by other protocols. Also, commands to control data acquisition devices are not included in this document.

This document does not put requirements on encryption or authentication of data. However, data is canonically encoded in the format of this document, which allows the application of data security algorithms.

This document does not replace IEC 62755. This document provides optional mechanisms to include IEC 62755 formatted data into list-mode data.

This document does not apply to data acquisition devices used as instrumentation and control systems in nuclear power plants, nor other safety-critical applications. It is also not intended for medical applications.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 62755, *Radiation protection instrumentation – Data format for radiation instruments used in the detection of illicit trafficking of radioactive materials*

ISO/IEC 8824-1:2015, ITU-T X.680, *Information technology – Abstract Syntax Notation One (ASN.1): Specification of basic notation*

ISO/IEC 8825-7, ITU-T X.696, *Information technology – ASN.1 encoding rules: Specification of Octet Encoding Rules (OER) and Technical Corrigendum 1 and Technical Corrigendum 2*