

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – Dose area product meters

Appareils électromédicaux – Radimètres de produit exposition-surface





THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – Dose area product meters

Appareils électromédicaux – Radimètres de produit exposition-surface

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-7591-7

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD	4
INTRODUCTION	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	8
4 General requirements	14
4.1 Performance requirements	14
4.2 Minimum EFFECTIVE RANGES of DOSE AREA PRODUCT and DOSE AREA PRODUCT RATE.....	14
4.3 Plane of measurement	14
4.4 REFERENCE VALUES and STANDARD TEST CONDITIONS.....	14
4.5 General test conditions	15
4.5.1 STANDARD TEST CONDITIONS	15
4.5.2 Test of components	15
4.5.3 STABILIZATION TIME.....	15
4.5.4 Adjustments during test	16
4.5.5 Uniformity of RADIATION field	16
4.6 Statistical fluctuations	16
4.7 Uncertainty of measurement	17
4.8 Constructional requirements as related to performance.....	17
4.8.1 Display	17
4.8.2 Indication of polarizing voltage failure.....	17
4.8.3 Over-ranging	17
4.8.4 Indication of reset or other inactive condition	18
4.8.5 RADIATION DETECTOR	18
4.9 STABILITY CHECK DEVICE	18
4.10 Adjustment.....	19
4.11 Electrical safety	20
5 Limits of PERFORMANCE CHARACTERISTICS under STANDARD TEST CONDITIONS	20
5.1 Classification of DOSE AREA PRODUCT METERS according to LIMITS OF VARIATION	20
5.1.1 REFERENCE-CLASS DOSE AREA PRODUCT METERS	20
5.1.2 FIELD-CLASS DOSE AREA PRODUCT METERS	20
5.2 LINEARITY	20
5.3 Warning function.....	20
5.4 Repeatability.....	21
5.5 RESOLUTION of reading	21
5.6 STABILIZATION TIME	21
5.7 Reset on DOSE AREA PRODUCT ranges	21
5.8 Drift of INDICATED VALUES.....	21
5.9 Long term stability	22
5.10 RESPONSE TIME.....	22
5.11 Spatial uniformity of RESPONSE.....	23
6 LIMITS OF VARIATION for effects of INFLUENCE QUANTITIES	23
6.1 General.....	23
6.2 Energy dependence of RESPONSE.....	23

6.3	DOSE AREA PRODUCT RATE dependence of DOSE AREA PRODUCT measurements.....	23
6.3.1	MEASURING ASSEMBLY.....	23
6.3.2	IONIZATION CHAMBER – Recombination losses	24
6.4	IRRADIATION TIME	24
6.5	Field size	24
6.6	Operating voltage	24
6.7	Air pressure	25
6.8	Temperature and humidity	25
6.9	Air density fluctuation in the IONIZATION CHAMBER	25
6.10	Electromagnetic compatibility.....	25
6.10.1	General	25
6.10.2	Electrostatic discharge	26
6.10.3	Radiated electromagnetic fields	26
6.10.4	Conducted disturbances induced by bursts and high frequencies.....	26
6.10.5	Surges.....	27
6.10.6	Voltage dips, short interruptions and voltage VARIATIONS	27
6.11	COMBINED STANDARD UNCERTAINTY	27
7	Marking	29
7.1	MEASURING ASSEMBLY	29
7.2	RADIATION DETECTOR	30
8	ACCOMPANYING DOCUMENTS	30
	Bibliography.....	32
	INDEX OF DEFINED TERMS.....	33
	Table 1 – Minimum EFFECTIVE RANGES – DOSE AREA PRODUCT	14
	Table 2 – Minimum EFFECTIVE RANGES – DOSE AREA PRODUCT RATE	14
	Table 3 – REFERENCE VALUES and STANDARD TEST CONDITIONS.....	15
	Table 4 – Number of readings required to detect true differences Δ (95 % confidence level) between two sets of instrument readings	16
	Table 5 – LIMITS OF VARIATION for the effects of INFLUENCE QUANTITIES	19
	Table 6 – Maximum values for the COEFFICIENT OF VARIATION, V_{\max}	21
	Table 7 – Climatic conditions	25
	Table 8 – Example for assessment of the COMBINED STANDARD UNCERTAINTY – FIELD-CLASS DOSE AREA PRODUCT METER.....	28
	Table 9 – Example for assessment of the COMBINED STANDARD UNCERTAINTY – REFERENCE-CLASS DOSE AREA PRODUCT METER	29

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT – DOSE AREA PRODUCT METERS

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60850 has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published 2000, and constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) a second class of devices is introduced with tighter uncertainty tolerances;
- b) this document has been expanded to include detectors other than ionization chambers;
- c) radiation qualities have been updated to the new definitions according to IEC 61267;
- d) a requirement on the linearity of the dose area product rate measurement was added;
- e) changed chamber light transmission requirement from 70 % to 60 %.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
62C/744/FDIS	62C/751/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- explanations, advice, general statements, exceptions and references: small roman type;
- *test specifications: italic type*;
- TERMS USED THROUGHOUT THIS STANDARD WHICH HAVE BEEN DEFINED IN CLAUSE 3 OR LISTED IN THE INDEX: SMALL CAPITALS.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

Diagnostic radiology is the largest contributor to man-made ionizing radiation to which the public is exposed. The reduction in the exposure received by PATIENTS undergoing MEDICAL RADIOLOGICAL EXAMINATIONS or procedures has therefore become a central issue in recent years. The purpose of routine measurement of DOSE AREA PRODUCT is to help in achieving an overall reduction in the radiation received by PATIENTS undergoing MEDICAL RADIOLOGICAL EXAMINATIONS. Provided adequate records are kept, it is possible to determine PATIENT doses, to compare different examination techniques, to establish a technique giving minimum RADIATION to a PATIENT, and to ensure a maintenance of that technique; in this respect, such measurements have a place of particular importance in training establishments. Examination of records may also indicate a deterioration in the efficiency of the image-production system.

MEDICAL ELECTRICAL EQUIPMENT – DOSE AREA PRODUCT METERS

1 Scope

This document specifies the performance and testing of DOSE AREA PRODUCT METERS intended to measure DOSE AREA PRODUCT and/or DOSE AREA PRODUCT RATE to which the PATIENT is exposed during MEDICAL RADIOLOGICAL EXAMINATIONS.

This document is applicable to the following types of DOSE AREA PRODUCT METERS:

- a) FIELD-CLASS DOSE AREA PRODUCT METERS normally used for the measurement of DOSE AREA PRODUCTS during MEDICAL RADIOLOGICAL EXAMINATIONS;
- b) REFERENCE-CLASS DOSE AREA PRODUCT METERS normally used for the CALIBRATION of FIELD-CLASS DOSIMETERS.

NOTE REFERENCE-CLASS DOSE AREA PRODUCT METERS can be used as FIELD-CLASS DOSE AREA PRODUCT METERS.

The object of this document is

- 1) to establish requirements for a satisfactory level of performance for DOSE AREA PRODUCT METERS, and
- 2) to standardize the methods for the determination of compliance with this level of performance.

Two levels of performance are specified:

- a lower level of performance applying to FIELD-CLASS DOSE AREA PRODUCT METERS;
- a higher level of performance applying to REFERENCE-CLASS DOSE AREA PRODUCT METERS.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60417, *Graphical symbols for use on equipment* (available at <http://www.graphical-symbols.info/equipment>)

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-2, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 62368-1, *Audio/video, information and communication technology equipment – Part 1: Safety requirements*