

American
National
Standard



ANSI/AAMI
ST67:2019

Sterilization of health care
products—Requirements
and guidance for selecting
a sterility assurance level
(SAL) for products labeled
“sterile”

Sterilization of health care products— Requirements and guidance for selecting a sterility assurance level (SAL) for products labeled “sterile”

Developed by
Association for the Advancement of Medical Instrumentation

Approved 29 July 2019 by
American National Standards Institute, Inc.

Abstract: This standard establishes requirements, guidance and a risk framework for identification of acceptable approach(es) for assurance of sterility and for selection of alternative SAL(s) or aseptic processing for health care product that cannot be terminally sterilized to achieve an SAL value of 10^{-6} .

Keywords: sterility assurance level (SAL), terminal sterilization, aseptic processing, biologics, combination devices

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties can obtain current information on all AAMI documents by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

AAMI
901 N Glebe Road, Suite 300
Arlington, VA 22203

© 2019 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of AAMI. Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI at 901 N Glebe Road, Suite 300, Arlington, VA 22203. Phone: (703) 525-4890; Fax: (703) 276-0793.

Printed in the United States of America

ISBN 978-1-57020-727-3

Contents

Page

Committee representation.....	iv
Foreword.....	vii
Introduction	viii
1 Scope	1
2 Normative references	1
3 Definitions.....	2
4 Determination of approach for assurance of sterility	4
5 Approaches to assure sterility other than use of an SAL value of 10^{-6}	8
Annexes	
A Background and historical application of sterility assurance.....	10
B Modification strategies to consider in order to achieve terminal sterilization at an SAL value of 10^{-6}	12
C Risk management considerations if an SAL value of 10^{-6} cannot be achieved	18
D Examples of risk management considerations to select an approach to assure sterility	23
Bibliography	30

Committee representation

Association for the Advancement of Medical Instrumentation

Assurance of Sterility Working Group

This standard was developed by the AAMI Assurance of Sterility Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the standard does not necessarily mean that all working group members votes for its approval. At the time this standard was published, the **AAMI Assurance of Sterility Working Group** had the following members:

Cochairs: Byron Lambert, PhD
Trabue Bryans

Members: Christopher Anderson, Johnson & Johnson
Ed Arscott, NAMSA
Jen Aleson, Quality, Microbiology & Sterilization Services LLC
Jenny Berg, Sterilucent Inc
Carolyn Braithwaite-Nelson, Spectranetics Corporation
David Brodersen, LexaMed Ltd
Trabue Bryans, BryKor LLC
Glenn Calvert, Tech Group North America
Tim Carlson, Becton Dickinson & Company
Emily Craven, Mevex Corporation
Sophia Czechowicz, Advanced Sterilization Products
Brian Drumheller, Femasys Inc
Gordon Ely, GEM Consulting
Dan Floyd, DuPont Tyvek Medical and Pharmaceutical Protection
Becky Gilsdorf, Healthmark
Zory Glaser, PhD, Johns Hopkins University-School of Public Health
Doug Harbrecht, Sterility Assurance LLC
Fatima Hasanian, Sotera Health
Jason Hedrick, Medtronic Inc Campus
Betty Howard, STERIS Corporation
Nupur Jain, Intuitive Surgical Inc
Valerie Kozik, Press Ganey/Soyring
Byron Lambert, PhD, Abbott Laboratories
Jack LeClair
David Lin, PhD, Taiwan Advanced Sterilization Technology Inc
Michelle Luebke, Baxter Healthcare Corporation
Tania Lupu, Case Medical Inc
Jeff Martin, Sterilization and Quality System Consulting LLC
Patrick McCormick, Bausch & Lomb Inc
Jami McLaren, Boston Scientific Corporation
Vanessa Molloy-Simard, TSO3 Inc
Gerry O'Dell, Gerry O'Dell Consulting
Jon Olson, Cantel Inc
Dave Parente, Ecolab
Claudia Romero-Waas, BSI Healthcare
Jody Rupert, WL Gore & Associates Inc
Manny Saavedra, Avanos Medical
Jessica Schoelerman, Accuratus Labs Services
Harry Shaffer, Sterilization Consulting Services
Myra Smith, FDA/CDRH
Michael Sprague, Ethide Laboratories Inc
Sopheak Srun, Quality Tech Services LLC
Liz Steiner, IUVO BioScience
Leslie Tavares, WuXi AppTec Inc
Don Tumminelli, HIGHPOWER Validation Testing & Lab Services Inc
Lisa Wakeman, Indiana University Health
Wendy Wangsgard, Nelson Laboratories LLC

Beverly Whitaker, Indigo Consulting Group LLC
Kristy Yates, 3M Healthcare
Roberto Zumbado, Philips

Alternates: Stacy Bohl, Boston Scientific Corporation
Brad Catalone, PhD, TSO3 Inc
Phil Cogdill, Medtronic Inc Campus
Chris Coker, Advanced Sterilization Products
Sean Colwell, WuXi AppTec Inc
Mary Ann Drosnock, Healthmark
Mark Duro, Cantel Inc
Kevin Grum, DuPont Tyvek Medical and Pharmaceutical Protection
Nichole Jackson, Ecolab
Satu King, Spectranetics Corporation
Mark Krocko, Mevex Corporation
Jim Kulla, LexaMed Ltd
Alison Laughlin, Baxter Healthcare Corporation
Gerry McDonnell, PhD, Johnson & Johnson
Nicole McLees, 3M Healthcare
Brian Nissen, Quality Tech Services LLC
Nicole Pasquino, Case Medical Inc
Janelle Pence, WL Gore & Associates
Kyrstan Polaski, STERIS Corporation
Jason Pope, Sotera Health
Nancy Rakiewicz, IUVO BioScience
Victoria Rodriguez, FDA/CDRH
Tyrone Rouse, Owens & Minor
Mike Schoene, Bausch & Lomb Inc
Mary Sheehan, BSI Healthcare
Tajinder Uppal, NAMSA
Hemant Vekaria, Abbott Laboratories
Brian Wallace, Intuitive Surgical Inc

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

Cochairs: Michael Scholla, PhD
Patrick Weixel, FDA/CDRH

Members: Anas Aljabo, SteriPro Canada Inc.
Brett Anderson, Cochlear Ltd
Richard Bancroft, Steris Corporation
Marie Brewer
Trabue Bryans, BryKor LLC
Jon Burdach, Nanosonics Limited
Tim Carlson, Becton Dickinson & Company
Phil Cogdill, Medtronic Inc Campus
Lena Cordie, Qualitas Professional Services LLC
Jackie Daley
Gordon Ely, GEM Consulting
Lisa Foster, Aduvo QS & SA Consulting
Dan Fowler, WuXi AppTec
Joel Gorski, NAMSA
Joyce Hansen, Johnson & Johnson
Stephanie Homuth
Gail Horvath, ECRI Institute
Clark Houghtling, Cosmed Group Inc

Sue Klacik, IAHCMM
Erin Kyle, Association of periOperative Registered Nurses
Byron Lambert, Abbott Laboratories
Michelle Luebke, Baxter Healthcare Corporation
Patrick McCormick, Bausch & Lomb Inc
Leslie Nichols, Mayo Clinic
Gerry O'Dell, Gerry O'Dell Consulting
David Pickard, BSI Healthcare
Adrian Ponce, Verrix LLC
Janet Prust, 3M Healthcare
Nancy Rakiewicz, IUVO BioScience
Michael Scholla, PhD, DuPont Tyvek Medical and Pharmaceutical Protection
Linda Schultz, Northside Hospital Surgical Services Atlanta
Kristen Singleton, Getinge USA
Joan Spear, B Braun of America Inc
Patrick Weixel, FDA/CDRH
Sid Wiggs
Stephen Yeadon, Boston Scientific Corporation
Bill Young, Sotera Health
Roberto Zumbado, Philips

Alternates: Stacy Bohl, Boston Scientific Corporation
Greg Crego, IUVO BioScience
Dan Floyd, DuPont Tyvek Medical and Pharmaceutical Protection
Bob Marrs, B Braun of America Inc
Gerry McDonnell, PhD, Johnson & Johnson
David McGoldrick, Abbott Laboratories
Kim Patton, Becton Dickinson & Company
Christine Render, Cosmed Group Inc
Mike Sadowski, Baxter Healthcare Corporation
Mark Smith, Getinge USA
Craig Wallace, 3M Healthcare
Lisa Ward, STERIS Corporation
Jill Warren, WuXi AppTec
Martell Winters, Sotera Health
Jon Wood, IAHCMM

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

Foreword

This standard was developed by the AAMI Assurance of Sterility Working Group under the auspices of the AAMI Sterilization Standards Committee.

The purpose of this standard, for product to be labelled as sterile, is to provide requirements, guidance and a risk framework for identification of acceptable approach(es) for assurance of sterility and for selection of alternative SAL(s) or aseptic processing for health care products that cannot be terminally sterilized to achieve an SAL value of 10^{-6} .

Compared with the previous edition of the standard, changes in ANSI/AAMI ST67:2019 include:

- addition of a section of alternate SAL;
- addition of a section on aseptic processing;
- inclusion of a new annex on modification strategies to achieve terminal sterilization at an SAL of 10^{-6} ;
- new tables for risk aspects to consider for individual patients; and
- new tables for product SAL requirements.

This standard reflects the conscientious efforts of health care professionals, in cooperation with medical device and equipment manufacturers, to develop recommendations for quality systems for the processing of medical devices. It is not intended that these recommendations be construed as universally applicable in all circumstances. Also, it is recognized that in many cases these recommendations might not be immediately achievable. Therefore, the document should be used to guide personnel towards desirable performance objectives, and all of its provisions should be considered and applied in the light of professional judgment and experience.

As used within the context of this document, “shall” indicates requirements strictly to be followed to conform to the standard. “Should” indicates that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the standard. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

The provisions of this standard should be reviewed routinely by departmental managers and quality management representatives and adapted to the needs of their particular institutions. Written policies, procedures, and work instructions should be developed and implemented in consultation with appropriate hospital committees (e.g., safety, infection prevention and control, and hazardous materials).

The concepts incorporated in this standard should be considered flexible and dynamic. The recommendations set forth in this document are reviewed and updated periodically to assimilate progressive technological developments. AAMI policies and procedures require that AAMI standards and recommended practices be reviewed and, if necessary, revised at least once every five years.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Technical Programs, AAMI, 901 N. Glebe Rd, Suite 300, Arlington, VA 22203.

NOTE—This foreword does not contain provisions of the AAMI Standard, *Sterilization of health care products—Requirements and guidance for selecting a sterility assurance level (SAL) for products labeled “sterile”* (AAMI ST67:2019), but it does provide important information about the development and intended use of the document. American National Standard ANSI/AAMI ST67:2019.

Introduction

Sterility (or the state of being sterile) is defined as being free of viable microorganisms. The sterility of a health care product can be achieved through validation and control of a terminal sterilization process (or processes), through aseptic processing (i.e., product components are separately sterilized prior to final aseptic assembly) when terminal sterilization is not feasible, or a combination of the two processes. Regardless of the process, there are other considerations to assuring that the outcome is a sterile product. Some of these considerations are product design, the manufacturing process, quality control of raw materials, process control and the method used to qualify the process. For health care products labeled 'sterile' attention to product and sterile barrier system or container characteristics, manufacturing facilities, controls, and other aspects of a quality system are required as defined in standards and government regulations.¹

The AAMI/ISO standards recognize that there are certain processes used in manufacturing for which the results cannot be fully verified by subsequent inspection and testing of the product. Examples of such processes are those that reduce the risk of microbial contamination, e.g., terminal sterilization and aseptic processing. For this reason, terminal sterilization and aseptic processing processes must be validated and monitored.

Healthcare products produced under standard manufacturing conditions in accordance with requirements for quality management systems typically have microorganisms on them and are nonsterile. The microbial load on product can be reduced by controlling bioburden of raw materials, personnel and the manufacturing environment in conjunction with:

- a) a terminal sterilization process; and/or
- b) a combination of component (e.g., raw materials, drug or biologic product, containers and/or devices) sterilization and aseptic processing.

During terminal sterilization processes, the manufactured product within its sterile barrier system or container is sterilized using a defined microbicidal process. The purpose of terminal sterilization is to achieve sterility by inactivating product bioburden using a validated process. Inactivation studies with microorganisms exposed to sterilizing agents (e.g., dry heat, moist heat, ethylene oxide, or radiation) approximates an exponential rate of kill. Mathematically there is always a finite probability that a microorganism might survive, regardless of the extent of treatment applied. For a given extent of treatment, the probability of survival is influenced by the number and resistance of microorganisms and the environment in which the microorganisms exist during treatment. Sterility in these cases is defined in terms of the probability of a viable microorganism on/in the product following sterilization. This mathematical probability is commonly referred to as a sterility assurance level (SAL).

The purpose of this standard, for product to be labelled as sterile, is to provide requirements, guidance and a risk framework for identification of acceptable approach(es) for assurance of sterility and for selection of alternative SAL(s) or aseptic processing for health care products that cannot be terminally sterilized to achieve an SAL value of 10^{-6} .

¹ For medical devices, refer to U.S. Food and Drug Administration's (FDA's) 21 CFR 820, International Organization for Standardization's (ISO's) ISO 13485, and European Medical Device Regulation. ANSI/AAMI/ISO 13485 is an application of the ISO 9000 series of quality management system standards.

For pharmaceutical products, refer to U.S. FDA's 21 CFR 210/211 and European Medical Device Regulation and biologics U.S. FDA's 21 CFR 600/601

Sterilization of health care products— Requirements and guidance for selecting a sterility assurance level (SAL) for products labeled “sterile”

1 Scope

1.1 Inclusions

1.1.1 This standard specifies requirements and provides guidance on selecting approaches to establish assurance of sterility for a health care product labelled as “sterile.” Approaches to assurance of sterility include terminal sterilization at the accepted SAL appropriate for the product (e.g., 10^{-6} or 10^{-3}), terminal sterilization at alternative SALs and aseptic processing.

1.1.2 This standard as it pertains to terminal sterilization applies to sterilization processes in which microorganisms are inactivated by physical and/or chemical means.

1.2 Exclusions

1.2.1 This standard does not address health care products that are not labeled “sterile.” For example, nonsterile health care products that possess antimicrobial properties or contain preservatives for the control of microbial levels are not addressed.

1.2.2 This standard is not applicable to the sterilization of used or reprocessed health care products.

1.2.3 This standard is not applicable to sterilization of health care products by filtration.

1.2.4 This standard does not specify requirements for development, validation and routine control of a process for inactivating the causative agents of spongiform encephalopathies such as scrapie, bovine spongiform encephalopathy and Creutzfeldt-Jakob disease. Specific recommendations have been produced in particular countries for the processing of materials potentially contaminated with these agents.

NOTE—See also ISO 22442-1, ISO 22442-2 and ISO 22442-3.

1.2.5 This standard does not attempt to define what is considered an acceptable level of risk to justify an alternative SAL. This risk assessment must be based on valid clinical and/or scientific data. As such, what is considered an acceptable level of risk might vary from product to product.

2 Normative references

The following normative references contain provisions that, through reference in the text, constitute provisions of this standard. For any dated reference, subsequent amendments to or revisions of the reference do not apply. However, parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent editions of the normative references indicated below.

ANSI/AAMI/ISO 14971, *Medical devices—Application of risk management to medical devices*

U.S. FOOD AND DRUG ADMINISTRATION. Current good manufacturing practice in manufacturing, processing, packing, or holding of drugs; general. Code of Federal Regulations, Title 21, Part 210/211.

U.S. FOOD AND DRUG ADMINISTRATION. Biological products: General. Code of Federal Regulations, Title 21, Part 600/601.