

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

---

**Electroacoustics – Hearing aids –  
Part 0: Measurement of the performance characteristics of hearing aids**

**Électroacoustique – Appareils de correction auditive –  
Partie 0: Mesure des caractéristiques fonctionnelles des appareils de correction  
auditive**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2022 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

---

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 300 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 19 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.



IEC 60118-0

Edition 4.0 2022-08

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

---

**Electroacoustics – Hearing aids –  
Part 0: Measurement of the performance characteristics of hearing aids**

**Électroacoustique – Appareils de correction auditive –  
Partie 0: Mesure des caractéristiques fonctionnelles des appareils de correction  
auditive**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

---

ICS 17.140.50

ISBN 978-2-8322-5591-9

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	7
1 Scope.....	9
2 Normative references .....	9
3 Terms and definitions .....	10
4 General conditions.....	15
4.1 Acoustic measurement method .....	15
4.2 Acoustic couplers and occluded ear simulator.....	15
4.3 Input signals and frequency range .....	17
4.4 Reporting of data .....	17
5 Test equipment.....	17
5.1 General.....	17
5.2 Test enclosure .....	18
5.2.1 General .....	18
5.2.2 Type 1 "anechoic-chamber" .....	18
5.2.3 Type 2 "test-box" .....	19
5.3 Measurement configuration .....	19
5.3.1 General .....	19
5.3.2 Measurement configuration for non-directional hearing aids .....	19
5.3.3 Measurement configuration for directional hearing aids .....	22
5.4 Sequential measurement method.....	23
5.5 Other measurement methods .....	24
5.6 Acoustic response measurement.....	24
5.7 Sound source system.....	24
5.8 Measurement system for the sound pressure level and distortion produced by a hearing aid .....	25
5.9 Electrical input .....	26
5.10 Direct-current measurement system.....	26
5.11 Magnetic field source for SPLIV measurements .....	26
5.12 Magnetic field source for hearing aids having induction pick-up coil for use with a telephone.....	27
6 Normal operating conditions for a hearing aid.....	27
6.1 General.....	27
6.2 Battery or supply voltage .....	28
6.3 Settings of controls .....	28
6.3.1 General .....	28
6.3.2 Full-on setting (FOS) .....	28
6.3.3 Reference test setting (RTS) .....	28
6.4 Ambient conditions.....	29
6.5 Acoustical connection to the hearing aid .....	29
6.6 Accessories .....	29
7 Basic acoustic hearing aid measurements .....	29
7.1 General.....	29
7.2 Frequency response curves .....	29
7.3 OSPL90 frequency response curve .....	30
7.4 Full-on gain frequency response curve.....	30
7.5 Basic frequency response curve .....	31
7.5.1 Test procedure .....	31

7.5.2	Frequency range of amplification .....	31
7.5.3	Reference test gain (RTG) .....	32
7.6	Total harmonic distortion .....	32
8	Special hearing aid measurements .....	33
8.1	General .....	33
8.2	Difference frequency distortion .....	33
8.2.1	Difference frequency distortion products .....	33
8.2.2	Total difference frequency distortion .....	35
8.3	Equivalent input noise (EIN) .....	36
8.4	Equivalent input noise in one-third-octave bands (EIN-one-third-octave) .....	37
8.5	Measurements of AGC hearing aids .....	39
8.5.1	General .....	39
8.5.2	Steady-state input-output characteristics .....	39
8.5.3	AGC characteristics (attack and release time) .....	40
8.6	Effects of tone control and gain control .....	40
8.6.1	Basic frequency response: effect of tone control .....	40
8.6.2	Frequency response: effect of gain control position .....	40
8.6.3	Characteristics of the gain control .....	41
8.7	Battery related measurements .....	41
8.7.1	General .....	41
8.7.2	Battery current measurement .....	41
8.7.3	Effects of variation of battery or supply voltage and internal resistance .....	41
9	Special measurements for hearing aids having non-acoustical inputs .....	42
9.1	General .....	42
9.2	Equivalence of output levels for a non-acoustical input .....	43
9.3	Measurement setup for "Wireless" input (WL) .....	43
9.3.1	General .....	43
9.3.2	Measurement setup .....	43
9.3.3	Wireless basic frequency response SPLWL .....	44
9.3.4	HFA-SPLWL .....	44
9.3.5	Frequency response with full-on setting .....	44
9.3.6	Equivalent wireless input sensitivity level (EWLS) .....	44
9.4	Hearing aids with electrical input .....	45
9.4.1	General .....	45
9.4.2	Basic frequency response (SPLEI) .....	45
9.4.3	HFA-SPLEI .....	45
9.4.4	Frequency response with full-on setting .....	45
9.4.5	Equivalent electrical input sensitivity level (EEIS) .....	45
9.4.6	Connector system for electrical input .....	45
9.5	Hearing aids having induction pick-up coil in a vertical magnetic field .....	46
9.5.1	General .....	46
9.5.2	Basic frequency response in a vertical magnetic field (SPLIV) .....	46
9.5.3	Frequency response with full-on setting (full-on SPLIV) .....	46
9.5.4	Effect of gain control position on frequency response .....	46
9.5.5	Harmonic distortion .....	47
9.5.6	HFA-SPLIV .....	47
9.5.7	Equivalent test loop sensitivity level (ETLS) .....	47
9.5.8	Full-on HFA-SPLIV .....	47
9.6	Hearing aids having induction pick-up coil for use with a telephone .....	48

9.6.1	SPLITS response curve .....	48
9.6.2	HFA-SPLITS .....	48
9.6.3	Relative simulated equivalent telephone sensitivity level (RSETS).....	49
10	Measurement of performance characteristics for production, supply and delivery quality assurance purposes .....	49
10.1	General.....	49
10.2	Nominal characteristics .....	49
10.3	Reference test gain.....	50
10.4	OSPL90 .....	50
10.5	Full-on gain.....	50
10.6	Frequency response curve.....	50
10.7	Bandwidth frequencies $f_1$ and $f_2$ .....	51
10.8	Battery or supply voltage .....	52
10.9	Battery current .....	52
10.10	Total harmonic distortion.....	52
10.11	Equivalent input noise level .....	52
10.12	Full-on HFA-SPLIV .....	53
10.13	Equivalent test loop sensitivity level (ETLS).....	53
10.14	Relative simulated equivalent telephone sensitivity level (RSETS).....	53
10.15	HFA-SPLITS .....	53
10.16	Equivalent electrical input sensitivity level (EEIS) .....	53
10.17	Steady-state input-output AGC characteristics .....	54
10.18	Nominal attack time and release time.....	54
11	Maximum permitted expanded uncertainty of measurements .....	54
Annex A	(normative) Simultaneous measurement method.....	57
A.1	General.....	57
A.2	Measurement method .....	57
A.3	Comparison of the simultaneous method.....	57
Annex B	(normative) Substitution measurement method .....	58
B.1	General.....	58
B.2	Measurement method .....	58
B.3	Comparison of the substitution method .....	59
Annex C	(normative) Effect of MLE on non-acoustic input hearing aids .....	60
C.1	Acoustic and non-acoustic input equivalence .....	60
C.2	In situ equivalence measurement in the case of a known MLE .....	61
C.3	HFA equivalence of output levels for non-acoustical inputs .....	62
C.3.1	General .....	62
C.3.2	In situ output level equivalency of wireless input sensitivity level (EWLS).....	63
C.3.3	In situ output level equivalency of electrical input sensitivity level (EEIS).....	63
C.3.4	In situ output level equivalency of test loop sensitivity level (ETLS) .....	63
C.3.5	In situ output level equivalency of telephone sensitivity level (RSETS) .....	64
Annex D	(informative) Examples of uncertainty calculation .....	65
D.1	General.....	65
D.2	Uncertainty calculations .....	65
D.3	Sound source system for test-box .....	65
D.4	Sound source system for anechoic-chamber .....	66
D.5	Hearing aid measurement .....	67

Annex E (informative) Comparison of 2 cm <sup>3</sup> coupler and 0,4 cm <sup>3</sup> coupler .....	68
E.1 General.....	68
E.2 Influence of sound source impedance .....	68
E.3 Comparison of frequency responses of the 0,4 cm <sup>3</sup> , the 2 cm <sup>3</sup> coupler and the occluded ear simulator .....	70
Bibliography.....	72
Figure 1 – Example of test arrangement for behind the ear hearing aid.....	21
Figure 2 – Example of test arrangement for in the ear hearing aid .....	22
Figure 3 – Example of test arrangement for directional hearing aid .....	23
Figure 4 – Telephone magnetic field simulator (TMFS) .....	27
Figure 5 – Example of OSPL90 curve and basic frequency response curve.....	30
Figure 6 – Example of determination of frequency range from basic frequency response curve .....	32
Figure 7 – Example of fundamental and difference frequency distortion products.....	35
Figure 8 – Example of total difference frequency distortion .....	36
Figure 9 – Example of hearing aid acoustic gain .....	38
Figure 10 – Example of hearing aid output noise and test equipment noise.....	38
Figure 11 – Example of hearing aid equivalent input noise and ambient noise .....	39
Figure 12 – Example of a steady-state input-output characteristic.....	40
Figure 13 – Measurement setup for wireless input .....	44
Figure 14 – Example of hearing aids on TMFS for SPLITS test .....	48
Figure 15 – Example of tolerance limits and determination of frequency range from basic frequency response curve.....	51
Figure 16 – Manufacturer acceptance interval (a) and purchaser acceptance interval (b) with tolerance and maximum permitted uncertainty $U_{\max}$ .....	56
Figure B.1 – Example of test arrangement for the substitution method .....	59
Figure C.1 – Acoustic a) and non-acoustic b) transmission paths.....	61
Figure C.2 – Example of SPLIV equivalence for ITC .....	62
Figure E.1 – Deviation from the normalized coupler volume ratio as function of the effective source volume $V_s$ .....	69
Figure E.2 – Magnitude frequency responses of acoustic impedance of the 2 cm <sup>3</sup> , the 0,4 cm <sup>3</sup> coupler and various hearing aid types .....	70
Figure E.3 – Comparative measurement of the 0,4 cm <sup>3</sup> coupler, the 2 cm <sup>3</sup> coupler and the occluded ear simulator frequency responses.....	71
Table 1 – Overview on the use of acoustic couplers and occluded ear simulator .....	17
Table 2 – Resistors and open circuit voltages for zinc-air battery simulators .....	28
Table 3 – Distortion test frequencies and input sound pressure levels.....	33
Table 4 – OSPL90: Acceptable deviation from the nominal values: .....	50
Table 5 – Full-on gain: Acceptable deviation from the nominal value.....	50
Table 6 – Tolerances and acceptance intervals of the frequency response .....	51
Table 7 – Battery current: Acceptable deviation from the nominal value .....	52
Table 8 – Total harmonic distortion: Acceptable deviation from the nominal value.....	52

Table 9 – Equivalent input noise level: Acceptable deviation from the nominal value ..... 52

Table 10 – Full-on HFA-SPLIV: Acceptable deviation from the nominal value..... 53

Table 11 – ETLs: Acceptable deviation from the nominal value ..... 53

Table 12 – RSETS: Acceptable deviation from the nominal value..... 53

Table 13 – HFA-SPLITS: Acceptable deviation from the nominal value ..... 53

Table 14 – EEIS: Acceptable deviation from the nominal value ..... 54

Table 15 – Acceptable deviation from the nominal values at input levels of 50 dB and 90 dB ..... 54

Table 16 – Values of  $U_{\max}$  for measurements ..... 55

Table C.1 – HFA-MLE for various styles of hearing aids..... 63

Table D.1 – Uncertainty sound source system for test-box..... 66

Table D.2 – Uncertainty sound source system for anechoic-chamber ..... 66

Table D.3 – Uncertainty hearing aid measurement..... 67

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**ELECTROACOUSTICS –  
HEARING AIDS –****Part 0: Measurement of the performance characteristics of hearing aids****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60118-0 has been prepared by technical committee 29: Electroacoustics. It is an International Standard.

This fourth edition merges and updates the methods previously described in IEC 60118-0:2015 and IEC 60118-7:2005. It cancels and replaces the third edition of IEC 60118-0 published in 2015. This edition constitutes a technical revision.

Measurements for quality control as described in IEC 60118-7:2005 can be found in Clause 10 of this document.

This edition includes the following significant technical changes with respect to previous editions:

- a) the default use of an acoustic coupler according to IEC 60318-5,
- b) addition of the optional use of an occluded ear simulator according to IEC 60318-4,

- c) addition of the optional use of an acoustic coupler according to IEC 60318-8 (new standard based on IEC TS 62886) when information about the response above 8 kHz is needed, or the optional use of the acoustic coupler according to IEC 60318-8 for deep insert hearing aids,
- d) the addition of measurements of the performance of hearing aids for production, supply and delivery quality assurance purposes,
- e) corrected and updated measurement configuration and methods, adding the use of a sequential measurement as preferred configuration,
- f) updated and expanded measurement procedures for the non-acoustic inputs of the hearing aid.

NOTE The substitution method described in Annex B has no relation to the substitution method described in IEC 60118-0:2015.

The text of this International Standard is based on the following documents:

Draft	Report on voting
29/1126/FDIS	29/1129/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/standardsdev/publications](http://www.iec.ch/standardsdev/publications).

A list of all parts in the IEC 60118 series, published under the general title *Electroacoustics – Hearing aids*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

# ELECTROACOUSTICS – HEARING AIDS –

## Part 0: Measurement of the performance characteristics of hearing aids

### 1 Scope

This part of IEC 60118 gives recommendations for the measurement of the performance characteristics of air conduction hearing aids measured with an acoustic coupler or occluded ear simulator.

This document is applicable to the measurement and evaluation of the electroacoustical characteristics of hearing aids, for example for type testing and manufacturer data sheets.

This document is also applicable for the measurement of the performance characteristics of hearing aids for production, supply and delivery quality-assurance purposes.

The measurement results obtained by the methods specified in this document will express the performance under conditions of the measurement and can deviate substantially from the performance of the hearing aid under actual conditions of use.

This document primarily uses an acoustic coupler according to IEC 60318-5 which is only intended for loading a hearing aid with specified acoustic impedance and is not intended to reproduce the sound pressure in a person's ear. For measurements reflecting the output level in the normal human ear the occluded ear simulator according to IEC 60318-4 can be used. For extended high-frequency measurements and for deep insert hearing aids, the acoustic coupler according to IEC 60318-8 can be used.

This document also covers measurement of hearing aids with non-acoustic inputs, such as wireless, inductive or electrical input.

This document does not cover the measurement of hearing aids for simulated in situ working conditions, for which IEC 60118-8 can be applied.

This document does not cover the measurement of hearing aids under typical user settings and using a speech-like signal, for which IEC 60118-15 can be applied.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60118-12, *Hearing aids – Part 12: Dimensions of electrical connector systems*

IEC 60318-4:2010, *Electroacoustics – Simulators of human head and ear – Part 4: Occluded-ear simulator for the measurement of earphones coupled to the ear by means of ear inserts*

IEC 60318-5, *Electroacoustics – Simulators of human head and ear – Part 5: 2 cm<sup>3</sup> coupler for the measurement of hearing aids and earphones coupled to the ear by means of ear inserts*