

**INTERNATIONAL  
STANDARD**

**IEC  
60601-2-47**

First edition  
2001-07

---

---

**Medical electrical equipment –**

**Part 2-47:  
Particular requirements for the safety,  
including essential performance,  
of ambulatory electrocardiographic systems**

*Appareils électromédicaux –*

*Partie 2-47:  
Règles particulières de sécurité et performances essentielles  
des systèmes d'électrocardiographie ambulatoires*



Reference number  
IEC 60601-2-47:2001(E)

## Publication numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series. For example, IEC 34-1 is now referred to as IEC 60034-1.

## Consolidated editions

The IEC is now publishing consolidated versions of its publications. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

## Further information on IEC publications

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology. Information relating to this publication, including its validity, is available in the IEC Catalogue of publications (see below) in addition to new editions, amendments and corrigenda. Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is also available from the following:

- **IEC Web Site** ([www.iec.ch](http://www.iec.ch))
- **Catalogue of IEC publications**  
The on-line catalogue on the IEC web site ([www.iec.ch/catlg-e.htm](http://www.iec.ch/catlg-e.htm)) enables you to search by a variety of criteria including text searches, technical committees and date of publication. On-line information is also available on recently issued publications, withdrawn and replaced publications, as well as corrigenda.
- **IEC Just Published**  
This summary of recently issued publications ([www.iec.ch/JP.htm](http://www.iec.ch/JP.htm)) is also available by email. Please contact the Customer Service Centre (see below) for further information.
- **Customer Service Centre**  
If you have any questions regarding this publication or need further assistance, please contact the Customer Service Centre:

Email: [custserv@iec.ch](mailto:custserv@iec.ch)  
Tel: +41 22 919 02 11  
Fax: +41 22 919 03 00

# INTERNATIONAL STANDARD

# IEC 60601-2-47

First edition  
2001-07

---

---

## Medical electrical equipment –

### **Part 2-47: Particular requirements for the safety, including essential performance, of ambulatory electrocardiographic systems**

#### *Appareils électromédicaux –*

#### *Partie 2-47: Règles particulières de sécurité et performances essentielles des systèmes d'électrocardiographie ambulatoires*

© IEC 2001 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission 3, rue de Varembé Geneva, Switzerland  
Telefax: +41 22 919 0300 e-mail: [inmail@iec.ch](mailto:inmail@iec.ch) IEC web site <http://www.iec.ch>



Commission Electrotechnique Internationale  
International Electrotechnical Commission  
Международная Электротехническая Комиссия

PRICE CODE

**S**

*For price, see current catalogue*

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6

### SECTION ONE – GENERAL

1 Scope and object .....	7
2 Terminology and definitions.....	8
5 Classification.....	10
6 Identification, marking and documents.....	10

### SECTION TWO – ENVIRONMENTAL CONDITIONS

10 Environmental conditions.....	11
----------------------------------	----

### SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS

20 Dielectric strength .....	12
------------------------------	----

### SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS

21 Mechanical strength .....	12
------------------------------	----

### SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION

36 Electromagnetic compatibility .....	13
--	----

### SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES

### SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS

### SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT

50 Accuracy of operating data.....	15
51 Protection against hazardous output.....	22

### SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS

### SECTION TEN – CONSTRUCTIONAL REQUIREMENTS

56 Components and general assembly .....	29
--	----

Appendix L (normative) References – Publications mentioned in this standard.....	35
Annex AA (informative) Guidance and rationale.....	36
Figure 101 – Test set-up for conductive emission test according 36.201.1. ....	30
Figure 102 – Test set-up for radiated emission and radiated immunity test according to 36.201.1 and 36.202.2.....	31
Figure 103 – Test signal for input dynamic range test according to 51.5.1.....	32
Figure 104 – General test circuit for 51.5.....	32
Figure 105 – Test circuit for common mode rejection according to 51.5.3.....	33
Figure 106 – Test circuit for pacemaker pulse tolerance according to 51.5.11 .....	34
Table 101 – LEAD colour codes.....	10
Table 102 – Reporting requirements for standard analyser outputs .....	16
Table 103 – Reporting requirements for optional analyser outputs .....	16
Table 104 – Beat-by-beat matrix.....	19
Index of defined terms.....	44

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-47: Particular requirements for the safety, including essential performance, of ambulatory electrocardiographic systems**

## FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a world-wide organisation for standardisation comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardisation in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees, any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organisations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organisation for Standardisation (ISO) in accordance with conditions determined by agreement between the two organisations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-47 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this standard is based upon the following documents:

FDIS	Report on voting
62D/408/FDIS	62D/411/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 3.

Annex AA is for information only.

In this Particular Standard the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type,
- notes, explanations, advice, introductions, general statements, exceptions and references: in smaller type,
- *test specifications: in italic type,*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD OR THIS PARTICULAR STANDARD: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until 2005. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

A bilingual edition of this publication may be issued at a later date.

## INTRODUCTION

This Particular Standard concerns the safety of AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS. It amends and supplements IEC 60601-1 (second edition 1988): *Medical electrical equipment – Part 1: General requirements for safety*, as amended by its amendment 1 (1991) and its amendment 2 (1995), hereinafter referred to as the General Standard. The requirements of this Particular Standard take priority over those of the General Standard.

A “General guidance and rationale” for the requirements of this Particular Standard is included in annex AA.

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this Standard.

An asterisk (\*) by a clause or subclause number indicates that some explanatory notes are given in annex AA of this Particular Standard.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-47: Particular requirements for the safety, including essential performance, of ambulatory electrocardiographic systems

#### SECTION ONE – GENERAL

The clauses and subclauses of this section of the General Standard apply except as follows:

#### 1 Scope and object

This clause of the General Standard applies except as follows:

##### 1.1 Scope

*Addition:*

This Particular Standard specifies the particular safety requirements for AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS, as defined in 2.101.

Within the scope of this standard are systems of the following types:

- a) systems that provide continuous recording and continuous analysis of the ECG allowing full re-analysis giving essentially similar results. The systems may first record and store the ECG and analyse it later on a separate unit, or record and analyse the ECG simultaneously. The type of storage media used is irrelevant with regard to this standard;
- b) systems that provide continuous analysis and only partial or limited recording not allowing a full re-analysis of the ECG.

The safety aspects of this standard apply to all types of systems falling in one of the above-mentioned categories.

If the ambulatory electrocardiographic system offers automatic ECG analysis, minimal performance requirements for measurement and analysis functions apply. Medical electrical equipment covered by IEC 60601-2-25 and IEC 60601-2-27 are excluded from the scope of this standard.

This standard does not apply to systems that do not continuously record and analyse the ECG (for example, 'intermittent event recorders').

##### 1.2 Object

*Replacement:*

The object of this Particular Standard is to establish particular requirements for the safety, including essential performance, of AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS.